

LB128215


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52252



A= 0.0000e+000

B= 1.9440e-004 slope

C= -4.9123e-002 y-intercept

Rho= 0.9991944

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/p
0.00	0.000	-0.037	-0.037	62	0.000	62					=
0.05	0.050					419	0				-28
0.20	0.200	0.144	-0.056	996	0.0 %	996					11
2.50	2.500	2.768	0.268	14492	0.0 %	14492					-2
5.00	5.000	4.898	-0.102	25450	0.0 %	25450					-2
7.50	7.500	7.318	-0.182	37897	0.0 %	37897					1
10.0	10.000	10.108	0.108	52252	0.0 %	52252					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	62	-		0 SFAM01.1	PPB	11/7/2023 13:44	Std
	0.2 S01	996	-		0.2 SFAM01.1	PPB	11/7/2023 13:46	Std
	2.5 S02	14492	-		2.5 SFAM01.1	PPB	11/7/2023 13:49	Std
	5 S03	25450	-		5 SFAM01.1	PPB	11/7/2023 13:53	Std
	7.5 S04	37897	-		7.5 SFAM01.1	PPB	11/7/2023 13:58	Std
	10 S05	52252	-		10 SFAM01.1	PPB	11/7/2023 14:03	Std
ICV060	ICV060	22127	4.2523 -		SFAM01.1	PPB	11/7/2023 14:06	SMPL
ICB060	ICB060	-242	-0.0962 -		SFAM01.1	PPB	11/7/2023 14:08	SMPL
CCV071	CCV071	25389	4.8864 -		SFAM01.1	PPB	11/7/2023 14:10	SMPL
CCB071	CCB071	-162	-0.0806 -		SFAM01.1	PPB	11/7/2023 14:13	SMPL
PB156899BL	PBS899	-90	-0.0666 -		SFAM01.1	PPB	11/7/2023 14:15	SMPL
O4967-01	MEZZ00	25092	4.8286 -		SFAM01.1	PPB	11/7/2023 14:17	SMPL
O4967-02	MEZZ01	130672	25.3529 -		SFAM01.1	PPB	11/7/2023 14:19	SMPL
O4967-03	MEZZ02	87501	16.9607 -		SFAM01.1	PPB	11/7/2023 14:22	SMPL
O4967-04	MEZZ10	24783	4.7686 -		SFAM01.1	PPB	11/7/2023 14:24	SMPL
O4967-05	MEZZ15	8801	1.6617 -		SFAM01.1	PPB	11/7/2023 14:26	SMPL
O4967-06	MEZZ16	18147	3.4786 -		SFAM01.1	PPB	11/7/2023 14:29	SMPL
O4967-07	MEZZ17	27356	5.2688 -		SFAM01.1	PPB	11/7/2023 14:31	SMPL
O4967-08	MEZZ18	31671	6.1076 -		SFAM01.1	PPB	11/7/2023 14:33	SMPL
O4967-09	MEZZ19	83639	16.2099 -		SFAM01.1	PPB	11/7/2023 14:36	SMPL
O4967-11	MEZY20	62024	12.008 -		SFAM01.1	PPB	11/7/2023 14:38	SMPL
O4967-12	MEZY25	60911	11.7917 -		SFAM01.1	PPB	11/7/2023 14:40	SMPL
O4967-13	MEZY26	66668	12.9108 -		SFAM01.1	PPB	11/7/2023 14:43	SMPL
O4967-14	MEZY27	145350	28.2062 -		SFAM01.1	PPB	11/7/2023 14:45	SMPL
O4967-15	MEZY27D	141576	27.4726 -		SFAM01.1	PPB	11/7/2023 14:47	SMPL
O4967-16	MEZY27S	140678	27.298 -		SFAM01.1	PPB	11/7/2023 14:50	SMPL
CCV072	CCV072	24480	4.7097 -		SFAM01.1	PPB	11/7/2023 14:52	SMPL
CCB072	CCB072	-340	-0.1152 -		SFAM01.1	PPB	11/7/2023 14:54	SMPL
O4967-02DLX5	MEZZ01	28732	5.5362 -		SFAM01.1	PPB	11/7/2023 14:56	SMPL
O4967-03DLX5	MEZZ02	19173	3.678 -		SFAM01.1	PPB	11/7/2023 14:59	SMPL
O4967-0*9DLX5	MEZZ19	18265	3.5015 -		SFAM01.1	PPB	11/7/2023 15:01	SMPL
O4967-11DLX2	MEZY20	32956	6.3574 -		SFAM01.1	PPB	11/7/2023 15:03	SMPL
O4967-12DLX2	MEZY25	31941	6.1601 -		SFAM01.1	PPB	11/7/2023 15:06	SMPL
O4967-13DLX2	MEZY26	34560	6.6692 -		SFAM01.1	PPB	11/7/2023 15:08	SMPL
O4967-14DLX5	MEZY27	30342	5.8492 -		SFAM01.1	PPB	11/7/2023 15:10	SMPL
O4967-15DLX5	MEZY27D	31294	6.0343 -		SFAM01.1	PPB	11/7/2023 15:13	SMPL
O4967-16DLX5	MEZY27S	31140	6.0043 -		SFAM01.1	PPB	11/7/2023 15:15	SMPL
CCV073	CCV073	25898	4.9853 -		SFAM01.1	PPB	11/7/2023 15:17	SMPL
CCB073	CCB073	-124	-0.0732 -		SFAM01.1	PPB	11/7/2023 15:19	SMPL