

LB128227


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

54585



A= 0.0000e+000

B= 1.8337e-004

C= 3.4802e-002

Rho= 0.9998863

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.050	0.050	83	0.000	83	0				0.1)
0.2	0.200	0.181	-0.019	798	0.0 %	798					1
2.5	2.500	2.447	-0.053	13157	0.0 %	13157					-10
5.0	5.000	5.062	0.062	27414	0.0 %	27414					-2
7.5	7.500	7.415	-0.085	40248	0.0 %	40248					-1
10.0	10.000	10.044	0.044	54585	0.0 %	54585					10

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	
0	83	-	0	7470A	PPB	11/8/2023 10:51	Std	
0.2	798	-	0.2	7470A	PPB	11/8/2023 10:53	Std	
2.5	13157	-	2.5	7470A	PPB	11/8/2023 10:55	Std	
5	27414	-	5	7470A	PPB	11/8/2023 10:57	Std	
7.5	40248	-	7.5	7470A	PPB	11/8/2023 11:02	Std	
10	54585	-	10	7470A	PPB	11/8/2023 11:08	Std	
ICV62	ICV62	21182	3.919	-	7470A	PPB	11/8/2023 11:11	SMPL
ICB62	ICB62	-271	-0.0149	-	7470A	PPB	11/8/2023 11:13	SMPL
CCV78	CCV78	27178	5.0186	-	7470A	PPB	11/8/2023 11:15	SMPL
CCB78	CCB78	-256	-0.0121	-	7470A	PPB	11/8/2023 11:17	SMPL
CRA	CRA	714	0.1657	-	7470A	PPB	11/8/2023 11:20	SMPL
HighStd	HighStd	53774	9.8956	-	7470A	PPB	11/8/2023 11:22	SMPL
ChkStd	ChkStd	36360	6.7023	-	7470A	PPB	11/8/2023 11:24	SMPL
PB156984E	PBW	-391	-0.0369	-	7470A	PPB	11/8/2023 11:27	SMPL
PB156984E	LCSW	19030	3.5244	-	7470A	PPB	11/8/2023 11:29	SMPL
O5102-01	67S-IDW-0	-272	-0.0151	-	7470A	PPB	11/8/2023 11:31	SMPL
O5102-01C	67S-IDW-0	-91	0.0181	-	7470A	PPB	11/8/2023 11:33	SMPL
O5102-01A	67S-IDW-0	21806	4.0335	-	7470A	PPB	11/8/2023 11:36	SMPL
O5102-01A	67S-IDW-0	20193	3.7377	-	7470A	PPB	11/8/2023 11:38	SMPL
O5256-04	WC-1	-345	-0.0285	-	7470A	PPB	11/8/2023 11:40	SMPL
CCV79	CCV79	25954	4.7941	-	7470A	PPB	11/8/2023 11:42	SMPL
CCB79	CCB79	-377	-0.0343	-	7470A	PPB	11/8/2023 11:45	SMPL
O5256-08	WC-11	-97	0.017	-	7470A	PPB	11/8/2023 11:47	SMPL
O5256-12	WC-10	-46	0.0264	-	7470A	PPB	11/8/2023 11:49	SMPL
O5257-04	WC-6	-32	0.0289	-	7470A	PPB	11/8/2023 11:52	SMPL
O5257-08	WC-2	-20	0.0311	-	7470A	PPB	11/8/2023 11:54	SMPL
O5257-12	WC-3	-12	0.0326	-	7470A	PPB	11/8/2023 11:56	SMPL
PB156940T	PB156940T	-7	0.0335	-	7470A	PPB	11/8/2023 11:58	SMPL
O5102-01LX5		47	0.0434	-	7470A	PPB	11/8/2023 12:01	SMPL
O5102-01A		21344	3.9487	-	7470A	PPB	11/8/2023 12:03	SMPL
CCV80	CCV80	26111	4.8229	-	7470A	PPB	11/8/2023 12:05	SMPL
CCB80	CCB80	-378	-0.0345	-	7470A	PPB	11/8/2023 12:07	SMPL