


LB128230

SFAM01.1

INSTRUMENT ID: CVJ



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/p</i>
0.00	0.000	0.082	0.082	83	0.000	83					
0.05	0.050					419	0				
0.20	0.200	0.211	0.011	824	0.0 %	824					
2.50	2.500	2.371	-0.129	13315	0.0 %	13315					
5.00	5.000	5.067	0.067	28899	0.0 %	28899					
7.50	7.500	7.399	-0.101	42380	0.0 %	42380					
10.0	10.000	10.070	0.070	57821	0.0 %	57821					

LB128230
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	83	-		0 SFAM01.1	PPB	11/8/2023 14:26	Std
	0.2 S01	824	-		0.2 SFAM01.1	PPB	11/8/2023 14:28	Std
	2.5 S02	13315	-		2.5 SFAM01.1	PPB	11/8/2023 14:31	Std
	5 S03	28899	-		5 SFAM01.1	PPB	11/8/2023 14:33	Std
	7.5 S04	42380	-		7.5 SFAM01.1	PPB	11/8/2023 14:44	Std
	10 S05	57821	-		10 SFAM01.1	PPB	11/8/2023 14:46	Std
ICV063	ICV063	21460	3.7802 -		SFAM01.1	PPB	11/8/2023 14:49	SMPL
ICB063	ICB063	-143	0.0432 -		SFAM01.1	PPB	11/8/2023 14:51	SMPL
CCV081	CCV081	28464	4.9917 -		SFAM01.1	PPB	11/8/2023 14:53	SMPL
CCB081	CCB081	-195	0.0343 -		SFAM01.1	PPB	11/8/2023 14:56	SMPL
PB157002BL	PBW002	57	0.0778 -		SFAM01.1	PPB	11/8/2023 14:58	SMPL
O5017-01	MA4970	-32	0.0624 -		SFAM01.1	PPB	11/8/2023 15:00	SMPL
O5017-04	MA4A79	32367	5.6669 -		SFAM01.1	PPB	11/8/2023 15:02	SMPL
O5017-05	MA49B9	-314	0.0137 -		SFAM01.1	PPB	11/8/2023 15:05	SMPL
O5017-06	MA49B9D	-56	0.0583 -		SFAM01.1	PPB	11/8/2023 15:07	SMPL
O5017-07	MA49B9S	5313	0.987 -		SFAM01.1	PPB	11/8/2023 15:09	SMPL
O5017-08	MA4919	-77	0.0547 -		SFAM01.1	PPB	11/8/2023 15:11	SMPL
O5017-09	MA49H5	-9	0.0664 -		SFAM01.1	PPB	11/8/2023 15:14	SMPL
O5017-10	MA4992	-14	0.0656 -		SFAM01.1	PPB	11/8/2023 15:16	SMPL
O5017-11	MA4915	-5	0.0671 -		SFAM01.1	PPB	11/8/2023 15:18	SMPL
O5017-12	MA49H7	4	0.0687 -		SFAM01.1	PPB	11/8/2023 15:21	SMPL
O5017-13	MA49B3	-10	0.0663 -		SFAM01.1	PPB	11/8/2023 15:23	SMPL
O5017-14	MA49R2	-4	0.0673 -		SFAM01.1	PPB	11/8/2023 15:25	SMPL
O5017-15	MA4911	3	0.0685 -		SFAM01.1	PPB	11/8/2023 15:27	SMPL
O5017-16	MA49H1	5	0.0688 -		SFAM01.1	PPB	11/8/2023 15:30	SMPL
O5017-17	MA49H2	-4	0.0673 -		SFAM01.1	PPB	11/8/2023 15:32	SMPL
O5017-18	MA49Q8	133	0.091 -		SFAM01.1	PPB	11/8/2023 15:34	SMPL
O5017-19	MA49A9	5	0.0688 -		SFAM01.1	PPB	11/8/2023 15:36	SMPL
O5017-20	MA49B1	4	0.0687 -		SFAM01.1	PPB	11/8/2023 15:39	SMPL
O5017-21	MA49Q4	13	0.0702 -		SFAM01.1	PPB	11/8/2023 15:41	SMPL
CCV082	CCV082	25984	4.5627 -		SFAM01.1	PPB	11/8/2023 15:43	SMPL
CCB082	CCB082	-249	0.0249 -		SFAM01.1	PPB	11/8/2023 15:45	SMPL
O5017-22	MA49P1	-53	0.0588 -		SFAM01.1	PPB	11/8/2023 15:48	SMPL
O5017-23	MA49P2	-26	0.0635 -		SFAM01.1	PPB	11/8/2023 15:50	SMPL
O5017-24	MA49N6	-28	0.0631 -		SFAM01.1	PPB	11/8/2023 15:52	SMPL
O5017-25	MA4991	0	0.068 -		SFAM01.1	PPB	11/8/2023 15:54	SMPL
PB157003BL	PBW003	0	0.068 -		SFAM01.1	PPB	11/8/2023 15:57	SMPL
O5018-01	MA4971	-3	0.0675 -		SFAM01.1	PPB	11/8/2023 15:59	SMPL
O5018-03	MA4A65	43077	7.5195 -		SFAM01.1	PPB	11/8/2023 16:01	SMPL
O5018-04	MA49A0	-446	-0.0092 -		SFAM01.1	PPB	11/8/2023 16:04	SMPL
O5018-05	MA49A0D	-129	0.0457 -		SFAM01.1	PPB	11/8/2023 16:06	SMPL
O5018-06	MA49A0S	4693	0.8798 -		SFAM01.1	PPB	11/8/2023 16:08	SMPL
O5018-07	MA4920	-106	0.0496 -		SFAM01.1	PPB	11/8/2023 16:10	SMPL
O5018-08	MA49H8	-42	0.0607 -		SFAM01.1	PPB	11/8/2023 16:13	SMPL
O5018-09	MA4993	26	0.0725 -		SFAM01.1	PPB	11/8/2023 16:15	SMPL

LB128230

INSTRUMENT ID : CV1

O5018-10	MA4916	21	0.0716 -	SFAM01.1	PPB	11/8/2023 16:17 SMPL
O5018-11	MA49H4	40	0.0749 -	SFAM01.1	PPB	11/8/2023 16:19 SMPL
O5018-12	MA49B4	42	0.0752 -	SFAM01.1	PPB	11/8/2023 16:22 SMPL
O5018-13	MA49R3	53	0.0772 -	SFAM01.1	PPB	11/8/2023 16:24 SMPL
O5018-14	MA4912	49	0.0765 -	SFAM01.1	PPB	11/8/2023 16:26 SMPL
O5018-15	MA49H6	57	0.0778 -	SFAM01.1	PPB	11/8/2023 16:28 SMPL
O5018-16	MA49H3	54	0.0773 -	SFAM01.1	PPB	11/8/2023 16:31 SMPL
CCV083	CCV083	25827	4.5356 -	SFAM01.1	PPB	11/8/2023 16:33 SMPL
CCB083	CCB083	-258	0.0234 -	SFAM01.1	PPB	11/8/2023 16:35 SMPL
O5018-17	MA49Q9	-11	0.0661 -	SFAM01.1	PPB	11/8/2023 16:38 SMPL
O5018-18	MA49B0	11	0.0699 -	SFAM01.1	PPB	11/8/2023 16:40 SMPL
O5018-19	MA49B2	85	0.0827 -	SFAM01.1	PPB	11/8/2023 16:42 SMPL
O5018-20	MA49Q5	95	0.0844 -	SFAM01.1	PPB	11/8/2023 16:44 SMPL
O5018-21	MA49Q0	61	0.0785 -	SFAM01.1	PPB	11/8/2023 16:47 SMPL
O5018-22	MA49P3	62	0.0787 -	SFAM01.1	PPB	11/8/2023 16:49 SMPL
O5018-23	MA49P0	72	0.0804 -	SFAM01.1	PPB	11/8/2023 16:51 SMPL
O5018-24	MA49A6	25	0.0723 -	SFAM01.1	PPB	11/8/2023 16:53 SMPL
CCV084	CCV084	25802	4.5313 -	SFAM01.1	PPB	11/8/2023 17:05 SMPL
CCB084	CCB084	-503	-0.019 -	SFAM01.1	PPB	11/8/2023 17:07 SMPL