



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128239

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/09/2023	08:40
2	CAL2	1	Water	NA	NA	20.3	7.00	11/09/2023	08:41
3	CAL3	1	Water	NA	NA	20.4	10.02	11/09/2023	08:44
4	ICV	1	Water	NA	NA	20.3	7.01	11/09/2023	08:45
5	CCV1	1	Water	NA	NA	20.6	2.01	11/09/2023	08:47
6	O5279-04	1	Solid	20.02	20	23.6	9.07	11/09/2023	08:55
7	O5279-04DUP	1	Solid	20.03	20	23.7	9.08	11/09/2023	08:56
8	O5279-08	1	Solid	20.02	20	23.4	8.64	11/09/2023	09:00
9	O5291-02	1	Solid	20.03	20	24.3	7.10	11/09/2023	09:05
10	O5292-02	1	Solid	20.04	20	24.7	8.48	11/09/2023	09:10
11	O5295-04	1	Solid	20.02	20	23.0	7.38	11/09/2023	09:15
12	O5311-03	1	Solid	20.03	20	24.0	9.14	11/09/2023	09:20
13	O5311-06	1	Solid	20.04	20	23.8	7.57	11/09/2023	09:25
14	O5311-09	1	Solid	20.02	20	23.7	7.51	11/09/2023	09:35
15	O5311-12	1	Solid	20.03	20	23.4	8.10	11/09/2023	09:40
16	CCV2	1	Water	NA	NA	20.6	12.02	11/09/2023	09:45
17	O5311-15	1	Solid	20.04	20	23.2	8.98	11/09/2023	09:50
18	O5311-18	1	Solid	20.02	20	23.2	8.35	11/09/2023	10:00
19	O5317-05	1	Solid	20.01	20	23.2	7.69	11/09/2023	10:10
20	O5317-06	1	Solid	20.02	20	20.7	4.25	11/09/2023	10:15

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	20.5	2.01	11/09/2023	10:20

WORKLIST(Hardcopy Internal Chain)

128239

WorkList Name : CORRSOVITY O5105

WorkList ID : 175373

Department : Wet-Chemistry

Date : 11-08-2023 08:14:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5291-02	QUEEN-PLAZA	Solid	Corrosivity	Cool 4 deg C	TULL02	L41	11/07/2023	9045D
O5292-02	CORONA	Solid	Corrosivity	Cool 4 deg C	TULL02	L41	11/07/2023	9045D
O5279-04	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	11/06/2023	9045D
O5279-08	MH-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	11/06/2023	9045D
O5295-04	TP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	11/07/2023	9045D
O5311-03	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	11/08/2023	9045D
O5311-06	TP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	11/08/2023	9045D
O5311-09	TP-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	11/08/2023	9045D
O5311-12	TP-4	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	11/08/2023	9045D
O5311-15	TP-5	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	11/08/2023	9045D
O5311-18	TP-6	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	11/08/2023	9045D
O5317-05	D3308	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	11/08/2023	9045D
O5317-06	D3317	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	11/08/2023	9045D

Date/Time 11/09/23 08:30

Raw Sample Received by: JB (WOC)

Raw Sample Relinquished by: [Signature]

Date/Time 11/09/23 13:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: JB (WOC)