



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB128241

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/09/2023	08:15
2	CAL2	1	Water	NA	NA	20.3	7.00	11/09/2023	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	11/09/2023	08:19
4	ICV	1	Water	NA	NA	20.5	7.01	11/09/2023	08:20
5	CCV1	1	Water	NA	NA	20.6	2.01	11/09/2023	08:25
6	O5280-02	1	Water	NA	NA	20.3	6.68	11/09/2023	08:30
7	O5280-02DUP	1	Water	NA	NA	20.4	6.69	11/09/2023	08:31
8	O5284-05	1	Water	NA	NA	20.9	9.22	11/09/2023	08:35
9	O5287-01	1	Water	NA	NA	20.1	5.52	11/09/2023	08:45
10	O5287-02	1	Water	NA	NA	20.7	5.10	11/09/2023	08:50
11	O5301-01	1	Water	NA	NA	20.1	5.45	11/09/2023	09:00
12	O5316-01	1	Water	NA	NA	23.4	7.02	11/09/2023	09:10
13	O5316-02	1	Water	NA	NA	24.5	6.97	11/09/2023	09:15
14	CCV2	1	Water	NA	NA	20.3	12.02	11/09/2023	09:22

12824)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o5301

WorkList ID : 175395

Department : Wet-Chemistry

Date : 11-08-2023 11:54:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5301-01	POOL-WATER	Water	pH	Cool 4 deg C	STER04	L31	11/07/2023	9040C
O5280-02	MH-1-GW	Water	pH	Cool 4 deg C	PSEG03	I41	11/06/2023	9040C
O5284-05	50167	Water	pH	Cool 4 deg C	PSEG03	L31	11/07/2023	9040C
O5287-01	50188	Water	pH	Cool 4 deg C	PSEG03	L11	11/07/2023	9040C
O5287-02	50131	Water	pH	Cool 4 deg C	PSEG03	L11	11/07/2023	9040C
O5316-01	MWB-1978	Water	pH	Cool 4 deg C	PSEG03	L31	11/08/2023	9040C
O5316-02	MWB-1979	Water	pH	Cool 4 deg C	PSEG03	L31	11/08/2023	9040C

Date/Time 11/09/23 08:10

Raw Sample Received by: 801201

Raw Sample Relinquished by: 801201

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: