



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128249

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	11/09/2023	09:45
2	O5047-05		1	6	47.00	116.6	116.6	11/09/2023	10:15
3	O5317-06		1	6	24.00	75.2	75.2	11/09/2023	10:45
4	O5317-06DUP		1	6	24.00	75.2	75.2	11/09/2023	11:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66128249

WorkList Name : fp-s-11-08

WorkList ID : 175376

Department : Wet-Chemistry

Date : 11-08-2023 08:25:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5047-05	PT-FP-SOIL	Solid	Flash Point	Cool 4 deg C	CHEM02	QA Of	10/23/2023	1010B
O5317-06	D3317	Solid	Flash Point	Cool 4 deg C	PSEG03	L31	11/08/2023	1010B

Date/Time 11/09/2023 09:40
 Raw Sample Received by: RM
 Raw Sample Relinquished by: RM

Date/Time 11/09/2023 11:20
 Raw Sample Received by: RM
 Raw Sample Relinquished by: RM