



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128250

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	11/09/2023	12:00
2	O5280-02		1	12	100.00	>212.0	>212.0	11/09/2023	12:30
3	O5282-01		1	14	100.00	>212.0	>212.0	11/09/2023	13:00
4	O5284-05		1	14	100.00	>212.0	>212.0	11/09/2023	13:30
5	O5287-01		1	12	100.00	>212.0	>212.0	11/09/2023	14:00
6	O5287-02		1	12	100.00	>212.0	>212.0	11/09/2023	14:30
7	O5287-02DUP		1	12	100.00	>212.0	>212.0	11/09/2023	15:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

628250

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-11-08

WorkList ID : 175375

Department : Wet-Chemistry

Date : 11-08-2023 08:25:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5280-02 R	MH-1-GW	Water	Flash Point	Cool 4 deg C	PSEG03	I41	11/06/2023	1010B
O5282-01 N	YORK-20231106	Water	Flash Point	Cool 4 deg C	AECO05	L31	11/06/2023	1010B
O5284-05 C	50167	Water	Flash Point	Cool 4 deg C	PSEG03	L31	11/07/2023	1010B
O5287-01 E	50188	Water	Flash Point	Cool 4 deg C	PSEG03	L11	11/07/2023	1010B
O5287-02 E	50131	Water	Flash Point	Cool 4 deg C	PSEG03	L11	11/07/2023	1010B

Date/Time 11/09/2023 11:50
 Raw Sample Received by: RMA
 Raw Sample Relinquished by: Jof Goe

Date/Time 11/09/2023 15:13
 Raw Sample Received by: RMA
 Raw Sample Relinquished by: RMA