


LB128260

7471B

INSTRUMENT NO: CV1

 μ Abs.:

59403



Linear

A= 0.0000e+000

B= 1.6862e-004

C= 8.7911e-003

Rho= 0.9996786

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/e)
0.0	0.000	0.026	0.026	103	0.000	103					
0.2	0.200	0.161	-0.039	900	0.0 %	900					-20
2.5	2.500	2.457	-0.043	14519	0.0 %	14519					-2
5.0	5.000	5.169	0.169	30603	0.0 %	30603					3
7.5	7.500	7.362	-0.138	43605	0.0 %	43605					-2
10.0	10.000	10.026	0.026	59403	0.0 %	59403					0

LB128260 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Nov 2023 06:43:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	103	-	0.0000	7471B	PPB	10 Nov 2023 06:48:48	S	Std
0.2 - 1	50.2	900	-	0.2000	7471B	PPB	10 Nov 2023 06:51:05	S	Std
2.5 - 1	52.5	14519	-	2.5000	7471B	PPB	10 Nov 2023 06:53:22	S	Std
5.0 - 1	55	30603	-	5.0000	7471B	PPB	10 Nov 2023 06:55:39	S	Std
7.5 - 1	57.5	43605	-	7.5000	7471B	PPB	10 Nov 2023 07:00:54	S	Std
10.0 - 1	60	59403	-	10.0000	7471B	PPB	10 Nov 2023 07:08:23	S	Std
ICV64 - 1	ICV64	23341	3.9446	-7471B	PPB	10 Nov 2023 07:11:07	U	SMPL	
ICB64 - 1	ICB64	-342	-0.0489	-7471B	PPB	10 Nov 2023 07:13:23	U	SMPL	
CCV85 - 1	CCV85	30637	5.1749	-7471B	PPB	10 Nov 2023 07:15:40	U	SMPL	
CCB85 - 1	CCB85	-362	-0.0523	-7471B	PPB	10 Nov 2023 07:17:56	U	SMPL	
CRA - 1	CRA	1140	0.2010	-7471B	PPB	10 Nov 2023 07:22:57	U	SMPL	
HighStd - 1	HighStd	59322	10.0118	-7471B	PPB	10 Nov 2023 07:25:11	U	SMPL	
ChkStd - 1	ChkStd	41660	7.0336	-7471B	PPB	10 Nov 2023 07:27:26	U	SMPL	
PB157031BL - 1	PBS	-473	-0.0710	-7471B	PPB	10 Nov 2023 07:29:44	U	SMPL	
PB157031BS - 1	LCSS	22121	3.7389	-7471B	PPB	10 Nov 2023 07:32:03	U	SMPL	
O5279-01 - 1	TP-1	10435	1.7684	-7471B	PPB	10 Nov 2023 07:34:20	U	SMPL	
O5279-05 - 1	MH-1	63814	10.7693	-7471B	PPB	10 Nov 2023 07:36:38	U	SMPL	
O5291-01 - 1	QUEEN-PLAZA	11473	1.9434	-7471B	PPB	10 Nov 2023 07:38:57	U	SMPL	
O5292-01 - 1	CORONA	8612	1.4610	-7471B	PPB	10 Nov 2023 07:41:16	U	SMPL	
O5293-01 - 1	SB-1	941	0.1675	-7471B	PPB	10 Nov 2023 07:43:33	U	SMPL	
CCV86 - 1	CCV86	30151	5.0929	-7471B	PPB	10 Nov 2023 07:45:49	U	SMPL	
CCB86 - 1	CCB86	-134	-0.0138	-7471B	PPB	10 Nov 2023 07:48:04	U	SMPL	
O5293-02 - 1	SB-2	1526	0.2661	-7471B	PPB	10 Nov 2023 07:50:20	U	SMPL	
O5293-03 - 1	SB-3	117	0.0285	-7471B	PPB	10 Nov 2023 07:52:36	U	SMPL	
O5293-04 - 1	SB-4	3988	0.6813	-7471B	PPB	10 Nov 2023 07:54:51	U	SMPL	
O5293-05 - 1	SB-5	2194	0.3787	-7471B	PPB	10 Nov 2023 07:57:07	U	SMPL	
O5293-06 - 1	SB-6	302	0.0597	-7471B	PPB	10 Nov 2023 07:59:22	U	SMPL	
O5293-07 - 1	SB-7	296	0.0587	-7471B	PPB	10 Nov 2023 08:01:38	U	SMPL	
O5293-08 - 1	SB-8	482	0.0901	-7471B	PPB	10 Nov 2023 08:03:55	U	SMPL	
O5295-01 - 1	TP-2	1315	0.2305	-7471B	PPB	10 Nov 2023 08:06:12	U	SMPL	
O5311-01 - 1	TP-1	3321	0.5688	-7471B	PPB	10 Nov 2023 08:08:28	U	SMPL	
O5311-04 - 1	TP-2	5066	0.8630	-7471B	PPB	10 Nov 2023 08:10:44	U	SMPL	
CCV87 - 1	CCV87	30260	5.1113	-7471B	PPB	10 Nov 2023 08:13:00	U	SMPL	
CCB87 - 1	CCB87	-117	-0.0109	-7471B	PPB	10 Nov 2023 08:15:15	U	SMPL	
O5311-07 - 1	TP-3	3290	0.5636	-7471B	PPB	10 Nov 2023 08:17:32	U	SMPL	
O5311-07DUP - 1	TP-3DUP	4173	0.7125	-7471B	PPB	10 Nov 2023 08:19:48	U	SMPL	
O5311-07MS - 1	TP-3MS	31046	5.2439	-7471B	PPB	10 Nov 2023 08:22:03	U	SMPL	
O5311-07MSD - 1	TP-3MSD	31357	5.2963	-7471B	PPB	10 Nov 2023 08:24:19	U	SMPL	
O5311-10 - 1	TP-4	10306	1.7466	-7471B	PPB	10 Nov 2023 08:26:36	U	SMPL	
O5311-13 - 1	TP-5	8968	1.5210	-7471B	PPB	10 Nov 2023 08:28:55	U	SMPL	
O5311-16 - 1	TP-6	12629	2.1383	-7471B	PPB	10 Nov 2023 08:31:12	U	SMPL	
PB157049BL - 1	PBS	30	0.0138	-7471B	PPB	10 Nov 2023 08:36:22	U	SMPL	
PB157049BS - 1	LCVSS	22891	3.8687	-7471B	PPB	10 Nov 2023 08:44:42	U	SMPL	
O5312-01 - 1	P-4(170FT)(5-10)	34996	5.9099	-7471B	PPB	10 Nov 2023 08:46:57	U	SMPL	
CCV88 - 1	CCV88	32453	5.4811	-7471B	PPB	10 Nov 2023 08:49:13	U	SMPL	
CCB88 - 1	CCB88	-158	-0.0179	-7471B	PPB	10 Nov 2023 08:51:31	U	SMPL	
O5312-01DUP - 1	P-4(170FT)(5-10)DUP	34865	5.8878	-7471B	PPB	10 Nov 2023 08:53:48	U	SMPL	
O5312-01MS - 1	P-4(170FT)(5-10)MS	50795	8.5740	-7471B	PPB	10 Nov 2023 08:56:03	U	SMPL	
O5312-01MSD - 1	P-4(170FT)(5-10)MSD	49014	8.2737	-7471B	PPB	10 Nov 2023 08:58:22	U	SMPL	
O5312-02 - 1	DUP-03	30200	5.1012	-7471B	PPB	10 Nov 2023 09:00:39	U	SMPL	
O5312-03 - 1	P-2(200FT)(5-10)	7029	1.1940	-7471B	PPB	10 Nov 2023 09:02:57	U	SMPL	
O5317-01 - 1	208	11496	1.9473	-7471B	PPB	10 Nov 2023 09:05:16	U	SMPL	
O5317-03 - 1	RB-21148	162551	27.4186	-7471B	PPB	10 Nov 2023 09:07:32	U	SMPL	
O5323-01 - 1	TP-1	3825	0.6538	-7471B	PPB	10 Nov 2023 09:09:49	U	SMPL	
CCV89 - 1	CCV89	29557	4.9928	-7471B	PPB	10 Nov 2023 09:12:11	U	SMPL	
CCB89 - 1	CCB89	-171	-0.0200	-7471B	PPB	10 Nov 2023 09:14:27	U	SMPL	
O5311-07LX5 - 1		680	0.1235	-7471B	PPB	10 Nov 2023 09:16:44	U	SMPL	
O5311-07A - 1		29236	4.9387	-7471B	PPB	10 Nov 2023 09:18:59	U	SMPL	
O5312-01LX5 - 1		7770	1.3190	-7471B	PPB	10 Nov 2023 09:21:15	U	SMPL	
O5312-01A - 1		56129	9.4734	-7471B	PPB	10 Nov 2023 09:23:32	U	SMPL	
O5356-01 - 1		1765	0.3064	-7471B	PPB	10 Nov 2023 09:25:48	U	SMPL	
O5279-05DLX2 - 1		29555	4.9924	-7471B	PPB	10 Nov 2023 09:28:05	U	SMPL	
O5317-03DLX5 - 1		37889	6.3977	-7471B	PPB	10 Nov 2023 09:30:20	U	SMPL	
CCV90 - 1	CCV90	28353	4.7896	-7471B	PPB	10 Nov 2023 09:32:38	U	SMPL	
CCB90 - 1	CCB90	-168	-0.0195	-7471B	PPB	10 Nov 2023 09:34:56	U	SMPL	