



Analytical Summary Report

Analysis Method: 9045D
Parameter: Corrosivity
Run Number: LB128264
BalanceID: WC SC-6

Analyst By : JIGNESH
Supervisor Review By : Iwona
Slope : 98.6
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/09/2023	17:20
2	CAL2	1	Water	NA	NA	20.3	7.00	11/09/2023	17:22
3	CAL3	1	Water	NA	NA	20.6	10.02	11/09/2023	17:25
4	ICV	1	Water	NA	NA	20.7	7.02	11/09/2023	17:30
5	CCV1	1	Water	NA	NA	20.7	2.01	11/09/2023	17:35
6	O5323-04	1	Solid	20.02	20	22.9	5.86	11/09/2023	17:40
7	O5323-04DUP	1	Solid	20.03	20	22.8	5.88	11/09/2023	17:41
8	O5351-02	1	Solid	20.05	20	21.4	7.09	11/09/2023	17:45
9	O5351-04	1	Solid	20.02	20	21.4	6.73	11/09/2023	17:47
10	CCV2	1	Water	NA	NA	20.5	12.03	11/09/2023	17:50

W 128264

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRSOVITY 053561

WorkList ID : 175461

Department : Wet-Chemistry

Date : 11-09-2023 17:01:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5323-04	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	I31	11/08/2023	9045D
O5351-02	VNJ-241	Solid	Corrosivity	Cool 4 deg C	PSEG03	L51	11/09/2023	9045D
O5351-04	72-11985	Solid	Corrosivity	Cool 4 deg C	PSEG03	L51	11/09/2023	9045D

Date/Time 11-09-23 17:15
Raw Sample Received by: 78 (WC)
Raw Sample Relinquished by: 50 (CSM)

Date/Time 11-09-23 18:30
Raw Sample Received by: 50 (CSM)
Raw Sample Relinquished by: 50 (WC)