

LB12826G


7470A

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

55825



A= 0.0000e+000

B= 1.8209e-004

C= -2.1560e-001

Rho= 0.9999165

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%w/w
0.0	0.000	-0.082	-0.082	735	0.000	735					
0.2	0.200	0.236	0.036	2480	0.0 %	2480	0				10
2.5	2.500	2.529	0.029	15073	0.0 %	15073					10
5.0	5.000	5.037	0.037	28847	0.0 %	28847					10
7.5	7.500	7.530	0.030	42537	0.0 %	42537					10
10.0	10.000	9.950	-0.050	55825	0.0 %	55825					10

LB128266 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Nov 2023 10:20:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	735	-	0.0000	7470A	PPB	10 Nov 2023 10:29:07	S	Std
0.2 - 1	50.2	2480	-	0.2000	7470A	PPB	10 Nov 2023 10:31:24	S	Std
2.5 - 1	52.5	15073	-	2.5000	7470A	PPB	10 Nov 2023 10:33:41	S	Std
5.0 - 1	55	28847	-	5.0000	7470A	PPB	10 Nov 2023 10:35:58	S	Std
7.5 - 1	57.5	42537	-	7.5000	7470A	PPB	10 Nov 2023 10:43:23	S	Std
10.0 - 1	50	55825	-	10.0000	7470A	PPB	10 Nov 2023 10:50:46	S	Std
ICV65 - 1	ICV65	21510	3.7012	-	7470A	PPB	10 Nov 2023 10:53:43	U	SMPL
ICB65 - 1	ICB65	1210	0.0047	-	7470A	PPB	10 Nov 2023 10:55:59	U	SMPL
CCV91 - 1	CCV91	27373	4.7688	-	7470A	PPB	10 Nov 2023 11:45:03	U	SMPL
CCB91 - 1	CCB91	734	-0.0819	-	7470A	PPB	10 Nov 2023 11:52:27	U	SMPL
CRA - 1	CRA	2181	0.1815	-	7470A	PPB	10 Nov 2023 12:06:14	U	SMPL
HighStd - 1	HighStd	51651	9.1895	-	7470A	PPB	10 Nov 2023 12:08:28	U	SMPL
ChkStd - 1	ChkStd	38871	6.8624	-	7470A	PPB	10 Nov 2023 12:10:44	U	SMPL
PB157030BL - 1	PBW	669	-0.0938	-	7470A	PPB	10 Nov 2023 12:15:09	U	SMPL
PB157030BS - 1	LCSW	22726	3.9226	-	7470A	PPB	10 Nov 2023 12:17:24	U	SMPL
O5279-04 - 1	TP-1	412	-0.1406	-	7470A	PPB	10 Nov 2023 12:19:40	U	SMPL
O5279-04DUP - 1	TP-1DUP	649	-0.0974	-	7470A	PPB	10 Nov 2023 12:21:58	U	SMPL
O5279-04MS - 1	TP-1MS	20199	3.4624	-	7470A	PPB	10 Nov 2023 12:24:15	U	SMPL
O5279-04MSD - 1	TP-1MSD	19485	3.3324	-	7470A	PPB	10 Nov 2023 12:26:32	U	SMPL
O5279-08 - 1	MH-1	432	-0.1369	-	7470A	PPB	10 Nov 2023 12:28:47	U	SMPL
CCV92 - 1	CCV92	29279	5.1158	-	7470A	PPB	10 Nov 2023 12:35:48	U	SMPL
CCB92 - 1	CCB92	672	-0.0932	-	7470A	PPB	10 Nov 2023 12:40:08	U	SMPL
O5284-02 - 1	SB-1	858	-0.0594	-	7470A	PPB	10 Nov 2023 12:42:23	U	SMPL
O5284-04 - 1	SB-2	895	-0.0526	-	7470A	PPB	10 Nov 2023 12:44:39	U	SMPL
O5291-02 - 1	QUEEN-PLAZA	719	-0.0847	-	7470A	PPB	10 Nov 2023 12:46:54	U	SMPL
O5292-02 - 1	CORONA	731	-0.0825	-	7470A	PPB	10 Nov 2023 12:49:10	U	SMPL
O5295-04 - 1	TP-2	845	-0.0617	-	7470A	PPB	10 Nov 2023 12:51:26	U	SMPL
O5311-03 - 1	TP-1	940	-0.0444	-	7470A	PPB	10 Nov 2023 12:53:42	U	SMPL
O5311-06 - 1	TP-2	975	-0.0381	-	7470A	PPB	10 Nov 2023 12:55:59	U	SMPL
O5311-09 - 1	TP-3	683	-0.0912	-	7470A	PPB	10 Nov 2023 12:58:16	U	SMPL
O5311-12 - 1	TP-4	767	-0.0759	-	7470A	PPB	10 Nov 2023 13:00:32	U	SMPL
O5311-15 - 1	TP-5	775	-0.0745	-	7470A	PPB	10 Nov 2023 13:02:48	U	SMPL
CCV93 - 1	CCV93	27190	4.7354	-	7470A	PPB	10 Nov 2023 13:05:04	U	SMPL
CCB93 - 1	CCB93	906	-0.0506	-	7470A	PPB	10 Nov 2023 13:07:19	U	SMPL
O5311-18 - 1	TP-6	679	-0.0920	-	7470A	PPB	10 Nov 2023 13:09:36	U	SMPL
O5317-05 - 1	D3308	662	-0.0951	-	7470A	PPB	10 Nov 2023 13:11:52	U	SMPL
O5317-06 - 1	D3317	1069	-0.0209	-	7470A	PPB	10 Nov 2023 13:14:08	U	SMPL
PB157052BL - 1	PBW	642	-0.0987	-	7470A	PPB	10 Nov 2023 13:16:23	U	SMPL
PB157052BS - 1	LCSW	19541	3.3426	-	7470A	PPB	10 Nov 2023 13:18:39	U	SMPL
O5280-02 - 1	MH-1-GW	19780	3.3861	-	7470A	PPB	10 Nov 2023 13:20:55	U	SMPL
O5280-02DUP - 1	MH-1-GWDUP	18544	3.1611	-	7470A	PPB	10 Nov 2023 13:23:12	U	SMPL
O5280-02MS - 1	MH-1-GWMS	34787	6.1188	-	7470A	PPB	10 Nov 2023 13:25:32	U	SMPL
O5280-02MSD - 1	MH-1-GWMSD	35345	6.2204	-	7470A	PPB	10 Nov 2023 13:27:51	U	SMPL
O5284-05DLX10 - 1	50167	484	-0.1275	-	7470A	PPB	10 Nov 2023 13:30:09	U	SMPL
CCV94 - 1	CCV94	26192	4.5537	-	7470A	PPB	10 Nov 2023 13:32:26	U	SMPL
CCB94 - 1	CCB94	916	-0.0488	-	7470A	PPB	10 Nov 2023 13:34:43	U	SMPL
O5302-02 - 1	COMP	1365	0.0330	-	7470A	PPB	10 Nov 2023 13:37:00	U	SMPL
O5316-01 - 1	NWB-1978	2500	0.2396	-	7470A	PPB	10 Nov 2023 13:39:16	U	SMPL
O5316-02 - 1	NWB-1979	454	-0.1329	-	7470A	PPB	10 Nov 2023 13:41:33	U	SMPL
PB156994TB - 1	PB156994TB	674	-0.0929	-	7470A	PPB	10 Nov 2023 13:43:51	U	SMPL
O5279-04LX5 - 1		133	-0.1914	-	7470A	PPB	10 Nov 2023 13:46:07	U	SMPL
O5279-04A - 1		18285	3.1139	-	7470A	PPB	10 Nov 2023 13:48:23	U	SMPL
O5280-02LX5 - 1		4471	0.5985	-	7470A	PPB	10 Nov 2023 13:50:40	U	SMPL
O5280-02A - 1		33963	5.9687	-	7470A	PPB	10 Nov 2023 13:52:58	U	SMPL
O5318-01 - 1		3027	0.3356	-	7470A	PPB	10 Nov 2023 13:55:14	U	SMPL
CCV95 - 1	CCV95	27709	4.8299	-	7470A	PPB	10 Nov 2023 13:57:31	U	SMPL
CCB95 - 1	CCB95	1053	-0.0239	-	7470A	PPB	10 Nov 2023 14:02:16	U	SMPL