



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128269

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	11/10/2023	11:00
2	CAL2	1	Water	NA	NA	20.5	7.00	11/10/2023	11:01
3	CAL3	1	Water	NA	NA	20.3	10.02	11/10/2023	11:05
4	ICV	1	Water	NA	NA	20.5	6.98	11/10/2023	11:10
5	CCV1	1	Water	NA	NA	20.5	2.01	11/10/2023	11:15
6	O5348-01	1	Water	NA	NA	20.3	4.40	11/10/2023	11:20
7	O5348-01DUP	1	Water	NA	NA	20.4	4.41	11/10/2023	11:21
8	O5354-03	1	Water	NA	NA	20.7	5.60	11/10/2023	11:25
9	CCV2	1	Water	NA	NA	20.5	12.02	11/10/2023	11:30

WORKLIST(Hardcopy Internal Chain)

128269

WorkList Name : PH O5354

WorkList ID : 175478

Department : Wet-Chemistry

Date : 11-10-2023 10:43:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5348-01	K. BP0W-3-20231108	Water	pH	Cool 4 deg C	TETR06	L31	11/08/2023	9040C
O5354-03	R OILY-WATER-TOTE	Water	pH	Cool 4 deg C	PSEG03	L41	11/09/2023	9040C

Date/Time 11-10-23 10:50

Raw Sample Received by: 30(WC)

Raw Sample Relinquished by: 12(WC)

Date/Time 11-10-23 12:30

Raw Sample Received by: 30(WC)

Raw Sample Relinquished by: 30(WC)