



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128272

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	11/10/2023	11:50
2	O5316-01		1	12	100.00	>212.0	>212.0	11/10/2023	12:20
3	O5316-02		1	14	100.00	>212.0	>212.0	11/10/2023	12:50
4	O5354-03		1	12	100.00	>212.0	>212.0	11/10/2023	13:20
5	O5354-03DUP		1	12	100.00	>212.0	>212.0	11/10/2023	13:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16128272

WorkList Name : fp-11-10

WorkList ID : 175465

Department : Wet-Chemistry

Date : 11-10-2023 07:33:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5316-01 F	MWB-1978	Water	Flash Point	Cool 4 deg C	PSEG03	L31	11/08/2023	1010B
O5316-02 F	MWB-1979	Water	Flash Point	Cool 4 deg C	PSEG03	L31	11/08/2023	1010B
O5354-03 C	OILY-WATER-TOTE	Water	Flash Point	Cool 4 deg C	PSEG03	L41	11/09/2023	1010B

Date/Time 11/10/2023 11:45
 Raw Sample Received by: RM [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 11/10/2023 14:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]