


LB128278

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

44449



Linear

A= 0.0000e+000

B= 2.2538e-004 slope

C= 2.1163e-002 y-intercept

Rho= 0.9998572

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.00	0.000	0.028	0.028	31	0.000	31					
0.05	0.050					419					
0.20	0.200	0.238	0.038	963	0.0 %	963	0				
2.50	2.500	2.394	-0.106	10527	0.0 %	10527					
5.00	5.000	5.065	0.065	22379	0.0 %	22379					
7.50	7.500	7.436	-0.064	32901	0.0 %	32901					
10.0	10.000	10.039	0.039	44449	0.0 %	44449					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	31	-		0 SFAM01.1	PPB	11/10/2023 15:58	Std
	0.2 S01	963	-		0.2 SFAM01.1	PPB	11/10/2023 16:03	Std
	2.5 S02	10527	-		2.5 SFAM01.1	PPB	11/10/2023 16:06	Std
	5 S03	22379	-		5 SFAM01.1	PPB	11/10/2023 16:08	Std
	7.5 S04	32901	-		7.5 SFAM01.1	PPB	11/10/2023 16:10	Std
	10 S05	44449	-		10 SFAM01.1	PPB	11/10/2023 16:13	Std
ICV066	ICV066	17104	3.876 -		SFAM01.1	PPB	11/10/2023 16:15	SMPL
ICB066	ICB066	-210	-0.0262 -		SFAM01.1	PPB	11/10/2023 16:18	SMPL
CCV001	CCV001	22227	5.0306 -		SFAM01.1	PPB	11/10/2023 16:20	SMPL
CCB001	CCB001	-215	-0.0273 -		SFAM01.1	PPB	11/10/2023 16:22	SMPL
PB157077BL	PBW077	346	0.0991 -		SFAM01.1	PPB	11/10/2023 16:25	SMPL
O5031-01	MA49A2	316	0.0924 -		SFAM01.1	PPB	11/10/2023 16:27	SMPL
O5031-05	MA4939	408	0.1131 -		SFAM01.1	PPB	11/10/2023 16:31	SMPL
O5031-06	MA4939D	412	0.114 -		SFAM01.1	PPB	11/10/2023 16:34	SMPL
O5031-07	MA4939S	5273	1.2096 -		SFAM01.1	PPB	11/10/2023 16:36	SMPL
O5031-08	MA49G3	343	0.0985 -		SFAM01.1	PPB	11/10/2023 16:38	SMPL
O5031-09	MA4903	380	0.1068 -		SFAM01.1	PPB	11/10/2023 16:40	SMPL
O5031-10	MA4907	395	0.1102 -		SFAM01.1	PPB	11/10/2023 16:43	SMPL
O5031-11	MA49G0	397	0.1106 -		SFAM01.1	PPB	11/10/2023 16:45	SMPL
O5031-12	MA4987	367	0.1039 -		SFAM01.1	PPB	11/10/2023 16:47	SMPL
O5031-13	MA48Z9	349	0.0998 -		SFAM01.1	PPB	11/10/2023 16:49	SMPL
O5031-14	MA49F8	366	0.1037 -		SFAM01.1	PPB	11/10/2023 16:52	SMPL
O5031-15	MA49X5	396	0.1104 -		SFAM01.1	PPB	11/10/2023 16:54	SMPL
O5031-16	MA49X1	387	0.1084 -		SFAM01.1	PPB	11/10/2023 16:56	SMPL
O5031-17	MA4985	394	0.11 -		SFAM01.1	PPB	11/10/2023 16:58	SMPL
O5031-18	MA49Y0	353	0.1007 -		SFAM01.1	PPB	11/10/2023 17:01	SMPL
O5031-19	MA49Z1	395	0.1102 -		SFAM01.1	PPB	11/10/2023 17:03	SMPL
O5031-20	MA49X6	356	0.1014 -		SFAM01.1	PPB	11/10/2023 17:05	SMPL
O5031-21	MA49J5	255	0.0786 -		SFAM01.1	PPB	11/10/2023 17:08	SMPL
CCV002	CCV002	19653	4.4505 -		SFAM01.1	PPB	11/10/2023 17:10	SMPL
CCB002	CCB002	-238	-0.0325 -		SFAM01.1	PPB	11/10/2023 17:15	SMPL
O5031-22	MA4979	305	0.0899 -		SFAM01.1	PPB	11/10/2023 17:17	SMPL
O5031-23	MA48Z7	296	0.0879 -		SFAM01.1	PPB	11/10/2023 17:20	SMPL
O5031-24	MA49X8	325	0.0944 -		SFAM01.1	PPB	11/10/2023 17:22	SMPL
O5031-25	MA49Y4	323	0.094 -		SFAM01.1	PPB	11/10/2023 17:24	SMPL
PB157078BL	PBW078	104	0.0446 -		SFAM01.1	PPB	11/10/2023 17:26	SMPL
O5032-01	MA49A3	286	0.0856 -		SFAM01.1	PPB	11/10/2023 17:29	SMPL
O5032-04	MA4940	326	0.0946 -		SFAM01.1	PPB	11/10/2023 17:33	SMPL
O5032-05	MA4940D	354	0.1009 -		SFAM01.1	PPB	11/10/2023 17:35	SMPL
O5032-06	MA4940S	5086	1.1674 -		SFAM01.1	PPB	11/10/2023 17:38	SMPL
O5032-07	MA49H0	310	0.091 -		SFAM01.1	PPB	11/10/2023 17:40	SMPL
O5032-08	MA4904	351	0.1003 -		SFAM01.1	PPB	11/10/2023 17:42	SMPL
O5032-09	MA4908	356	0.1014 -		SFAM01.1	PPB	11/10/2023 17:45	SMPL
O5032-10	MA49G1	347	0.0994 -		SFAM01.1	PPB	11/10/2023 17:47	SMPL
O5032-11	MA4988	349	0.0998 -		SFAM01.1	PPB	11/10/2023 17:49	SMPL
O5032-12	MA4900	362	0.1027 -		SFAM01.1	PPB	11/10/2023 17:51	SMPL

LB128278

INSTRUMENT ID : CV1

O5032-13	MA49F9	345	0.0989 -	SFAM01.1	PPB	11/10/2023 17:54 SMPL
O5032-14	MA49Y1	369	0.1043 -	SFAM01.1	PPB	11/10/2023 17:56 SMPL
O5032-15	MA49X9	388	0.1086 -	SFAM01.1	PPB	11/10/2023 17:58 SMPL
O5032-16	MA4986	347	0.0994 -	SFAM01.1	PPB	11/10/2023 18:00 SMPL
CCV003	CCV003	20598	4.6634 -	SFAM01.1	PPB	11/10/2023 18:03 SMPL
CCB003	CCB03	-173	-0.0178 -	SFAM01.1	PPB	11/10/2023 18:05 SMPL
O5032-17	MA49X4	325	0.0944 -	SFAM01.1	PPB	11/10/2023 18:07 SMPL
O5032-18	MA49Y3	327	0.0949 -	SFAM01.1	PPB	11/10/2023 18:09 SMPL
O5032-19	MA49X7	376	0.1059 -	SFAM01.1	PPB	11/10/2023 18:12 SMPL
O5032-20	MA49J6	368	0.1041 -	SFAM01.1	PPB	11/10/2023 18:14 SMPL
O5032-21	MA4980	368	0.1041 -	SFAM01.1	PPB	11/10/2023 18:16 SMPL
O5032-22	MA48Z8	359	0.1021 -	SFAM01.1	PPB	11/10/2023 18:19 SMPL
O5032-23	MA49Y2	343	0.0985 -	SFAM01.1	PPB	11/10/2023 18:21 SMPL
O5032-24	MA49Z2	379	0.1066 -	SFAM01.1	PPB	11/10/2023 18:23 SMPL
O5031-04	MA4A80	23755	5.375 -	SFAM01.1	PPB	11/10/2023 18:28 SMPL
O5032-03	MA4A66	34255	7.7414 -	SFAM01.1	PPB	11/10/2023 18:34 SMPL
CCV004	CCV004	21400	4.8442 -	SFAM01.1	PPB	11/10/2023 18:39 SMPL
CCB004	CCB004	-366	-0.0613 -	SFAM01.1	PPB	11/10/2023 18:41 SMPL