



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 11/13/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 11/11/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/12/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB128285

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5261-01	DCKV0	1	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
O5261-02	DCKV2	2	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
O5261-03	DCKV4	3	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
O5261-04	DCKR5	4	1.12	8.78	9.9	7.73	75.3	
O5261-05	DCKQ2	5	1.13	8.48	9.61	5.65	53.3	
O5261-06	DCKQ5	6	1.15	8.45	9.6	2.3	13.6	
O5261-07	DCKQ6	7	1.14	8.63	9.77	9.18	93.2	
O5261-08	DCKQ7	8	1.19	8.79	9.98	9.54	95.0	
O5261-09	DCKQ8	9	1.18	8.47	9.65	9.03	92.7	
O5261-10	DCKQ9	10	1.19	8.78	9.97	9.37	93.2	
O5261-11	DCKR0	11	1.11	8.77	9.88	9.15	91.7	
O5261-12	DCKR1	12	1.16	8.83	9.99	9.58	95.4	
O5261-13	DCKR6	13	1.16	8.80	9.96	4.08	33.2	
O5261-14	DCKR7	14	1.19	8.57	9.76	4.16	34.7	
O5261-15	DCKR8	15	1.12	8.74	9.86	2.32	13.7	
O5261-16	DCKT0	16	1.17	8.71	9.88	2.89	19.7	
O5261-17	DCKT5	17	1.11	8.81	9.92	8.43	83.1	
O5261-18	DCKT5MS	18	1.11	8.81	9.92	8.43	83.1	
O5261-19	DCKT5MSD	19	1.11	8.81	9.92	8.43	83.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$