



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Rubina

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB128290

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP105273
Turbidity Calibration std, 0NTU	WP105272
Turbidity Calibration std, 20NTU	WP105268
Turbidity Calibration std, 40NTU	WP105267
Turbidity Calibration std, 1NTU	WP105271
Turbidity Calibration std, 5NTU	WP105270
Turbidity Calibration - CCV std, 10 NTU	WP105269
10 NTU Standard 500 ml	W2971

Intercept: 0.1515

Slope: 1.074

Regression: 0.9996

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.162	0.01		11/13/2023	09:30
2	CAL2	1	1	1.260	1.03	3.2	11/13/2023	09:33
3	CAL3	5	1	5.510	4.99	-0.2	11/13/2023	09:36
4	CAL4	10	1	11.400	10.47	4.7	11/13/2023	09:40
5	CAL5	20	1	20.800	19.23	-3.9	11/13/2023	09:43
6	CAL6	40	1	43.400	40.27	0.7	11/13/2023	09:46



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Rubina

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB128290

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.800	9.915	11/13/2023	09:50
2	ICB		1	0.158	0.006	11/13/2023	09:53
3	CCV1	10	1	11.200	10.287	11/13/2023	09:56
4	CCB1		1	0.160	0.008	11/13/2023	10:00
5	LB128290BL		1	0.164	0.012	11/13/2023	10:04
6	O4699-09		1	0.784	0.589	11/13/2023	10:07
7	CCV2	10	1	11.400	10.473	11/13/2023	10:10
8	CCB2		1	0.162	0.010	11/13/2023	10:13