

LB128297


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51512



A= 0.0000e+000

B= 1.9335e-004

C= -3.8563e-002

Rho= 0.9998807

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	01011
0.0	0.000	-0.006	-0.006	171	0.000	171	0				14
0.2	0.200	0.173	-0.027	1095	0.0 %	1095					1
2.5	2.500	2.521	0.021	13239	0.0 %	13239					0
5.0	5.000	4.982	-0.018	25964	0.0 %	25964					1
7.5	7.500	7.608	0.108	39550	0.0 %	39550					1
10.0	10.000	9.921	-0.079	51512	0.0 %	51512					1

LB128297 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Nov 2023 13:30:16

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	171	-	0.00007470A	PPB		13 Nov 2023 13:43:00	S	Std
0.2 - 1	50.2	1095	-	0.20007470A	PPB		13 Nov 2023 13:45:17	S	Std
2.5 - 1	52.5	13239	-	2.50007470A	PPB		13 Nov 2023 13:47:35	S	Std
5.0 - 1	55	25964	-	5.00007470A	PPB		13 Nov 2023 13:49:51	S	Std
7.5 - 1	59.5	39550	-	7.50007470A	PPB		13 Nov 2023 13:52:07	S	Std
10.0 - 1	54	51512	-	10.00007470A	PPB		13 Nov 2023 13:54:23	S	Std
ICV70 - 1	ICV70	22004	4.2159	-7470A	PPB		13 Nov 2023 13:57:15	U	SMPL
ICB70 - 1	ICB70	-113	-0.0604	-7470A	PPB		13 Nov 2023 13:59:29	U	SMPL
CCV13 - 1	CCV13	25234	4.8404	-7470A	PPB		13 Nov 2023 14:01:45	U	SMPL
CCB13 - 1	CCB13	-105	-0.0589	-7470A	PPB		13 Nov 2023 14:04:00	U	SMPL
CRA - 1	CRA	1352	0.2228	-7470A	PPB		13 Nov 2023 14:08:45	U	SMPL
HighStd - 1	HighStd	50688	9.7619	-7470A	PPB		13 Nov 2023 14:11:00	U	SMPL
ChkStd - 1	ChkStd	36221	6.9647	-7470A	PPB		13 Nov 2023 14:13:15	U	SMPL
PB157083BL - 1	PBW	-13	-0.0411	-7470A	PPB		13 Nov 2023 14:18:24	U	SMPL
PB157083BS - 1	LCSW	19856	3.8006	-7470A	PPB		13 Nov 2023 14:20:39	U	SMPL
O5323-04 - 1	TP-1	-207	-0.0786	-7470A	PPB		13 Nov 2023 14:22:55	U	SMPL
O5323-04DUP - 1	TP-1DUP	-45	-0.0473	-7470A	PPB		13 Nov 2023 14:25:12	U	SMPL
O5323-04MS - 1	TP-1MS	22112	4.2368	-7470A	PPB		13 Nov 2023 14:27:29	U	SMPL
O5323-04MSD - 1	TP-1MSD	19505	3.7327	-7470A	PPB		13 Nov 2023 14:29:46	U	SMPL
O5345-02 - 1	NEW-PARKING-LOT-STOCKPILE-A	-221	-0.0813	-7470A	PPB		13 Nov 2023 14:32:02	U	SMPL
CCV14 - 1	CCV14	26201	5.0274	-7470A	PPB		13 Nov 2023 14:34:18	U	SMPL
CCB14 - 1	CCB14	-170	-0.0714	-7470A	PPB		13 Nov 2023 14:36:34	U	SMPL
O5345-09 - 1	NEW-PARKING-LOT-STOCKPILE-B	-45	-0.0473	-7470A	PPB		13 Nov 2023 14:38:51	U	SMPL
O5351-02 - 1	VNJ-241	13	-0.0360	-7470A	PPB		13 Nov 2023 14:41:07	U	SMPL
O5351-04 - 1	72-11985	19	-0.0349	-7470A	PPB		13 Nov 2023 14:43:22	U	SMPL
O5352-01 - 1	GREEN-1	41	-0.0306	-7470A	PPB		13 Nov 2023 14:45:38	U	SMPL
O5352-02 - 1	GREEN-2	52	-0.0285	-7470A	PPB		13 Nov 2023 14:47:54	U	SMPL
O5352-03 - 1	GREEN-3	42	-0.0304	-7470A	PPB		13 Nov 2023 14:50:10	U	SMPL
O5352-04 - 1	GREEN-4	111	-0.0171	-7470A	PPB		13 Nov 2023 14:52:26	U	SMPL
O5352-06 - 1	WHITE-1	137	-0.0121	-7470A	PPB		13 Nov 2023 14:54:43	U	SMPL
O5352-07 - 1	WHITE-2	85	-0.0221	-7470A	PPB		13 Nov 2023 14:57:00	U	SMPL
O5352-08 - 1	WHITE-3	187	-0.0024	-7470A	PPB		13 Nov 2023 14:59:17	U	SMPL
CCV15 - 1	CCV15	26125	5.0127	-7470A	PPB		13 Nov 2023 15:01:32	U	SMPL
CCB15 - 1	CCB15	-99	-0.0577	-7470A	PPB		13 Nov 2023 15:03:50	U	SMPL
O5352-09 - 1	WHITE-4	59	-0.0272	-7470A	PPB		13 Nov 2023 15:06:09	U	SMPL
O5352-10 - 1	YELLOW	69	-0.0252	-7470A	PPB		13 Nov 2023 15:08:25	U	SMPL
PB157084BL - 1	PBW	49	-0.0291	-7470A	PPB		13 Nov 2023 15:10:40	U	SMPL
PB157084BS - 1	LCSW	22727	4.3557	-7470A	PPB		13 Nov 2023 15:12:56	U	SMPL
O5354-03 - 1	OILY-WATER-TOTE	-186	-0.0745	-7470A	PPB		13 Nov 2023 15:15:11	U	SMPL
O5354-03DUP - 1	OILY-WATER-TOTEDUP	-61	-0.0504	-7470A	PPB		13 Nov 2023 15:17:28	U	SMPL
O5354-03MS - 1	OILY-WATER-TOTEMS	21530	4.1242	-7470A	PPB		13 Nov 2023 15:19:44	U	SMPL
O5354-03MSD - 1	OILY-WATER-TOTEMSD	22153	4.2447	-7470A	PPB		13 Nov 2023 15:22:01	U	SMPL
CCV16 - 1	CCV16	25568	4.9050	-7470A	PPB		13 Nov 2023 15:24:16	U	SMPL
CCB16 - 1	CCB16	-202	-0.0776	-7470A	PPB		13 Nov 2023 15:26:34	U	SMPL
PB157050TB - 1	PB157050TB	-83	-0.0546	-7470A	PPB		13 Nov 2023 15:28:52	U	SMPL
PB157069TB - 1	PB157069TB	-38	-0.0459	-7470A	PPB		13 Nov 2023 15:31:07	U	SMPL
O5323-04LX5 - 1		36	-0.0316	-7470A	PPB		13 Nov 2023 15:33:23	U	SMPL
O5323-04A - 1		21772	4.1710	-7470A	PPB		13 Nov 2023 15:35:38	U	SMPL
O5354-03LX5 - 1		-99	-0.0577	-7470A	PPB		13 Nov 2023 15:37:54	U	SMPL
O5354-03A - 1		21610	4.1397	-7470A	PPB		13 Nov 2023 15:40:11	U	SMPL
CCV17 - 1	CCV17	26038	4.9959	-7470A	PPB		13 Nov 2023 15:42:27	U	SMPL
CCB17 - 1	CCB17	-213	-0.0797	-7470A	PPB		13 Nov 2023 15:44:44	U	SMPL