

LB128307

245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46189

A= 0.0000e+000

B= 2.1588e-004

C= -6.8505e-003

Rho= 0.9999221

Accept=Accepted

Max. Conc.: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.010	0.010	80	0.000	80					110
0.05	0.050	0.062	0.012	319	0.0 %	319	0				24
0.2	0.200	0.200	-0.000	956	0.0 %	956					0
2.5	2.500	2.415	-0.085	11217	0.0 %	11217					3
5.0	5.000	5.075	0.075	23542	0.0 %	23542					2
7.5	7.500	7.523	0.023	34882	0.0 %	34882					0
10.0	10.000	9.964	-0.036	46189	0.0 %	46189					0

LB128367

LB128313

INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 14 Nov 2023 08:38:44

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	80	-	0.0000	245.1	PPB	14 Nov 2023 08:45:18	S	Std
0.05 - 1	50.85	319	-	0.0500	245.1	PPB	14 Nov 2023 08:47:37	S	Std
0.2 - 1	52.20	956	-	0.2000	245.1	PPB	14 Nov 2023 08:49:54	S	Std
2.5 - 1	52.5	11217	-	2.5000	245.1	PPB	14 Nov 2023 08:52:10	S	Std
5.0 - 1	55	23542	-	5.0000	245.1	PPB	14 Nov 2023 08:54:27	S	Std
7.5 - 1	57.5	34882	-	7.5000	245.1	PPB	14 Nov 2023 08:56:42	S	Std
10.0 - 1	61.0	46189	-	10.0000	245.1	PPB	14 Nov 2023 09:02:39	S	Std
ICV71 - 1	ICV71	18073	3.8948	-245.1		PPB	14 Nov 2023 09:07:24	U	SMPL
ICB71 - 1	ICB71	-277	-0.0666	-245.1		PPB	14 Nov 2023 09:09:39	U	SMPL
CCV17 - 1	CCV17	23813	5.1339	-245.1		PPB	14 Nov 2023 09:11:55	U	SMPL
CCB17 - 1	CCB17	-327	-0.0774	-245.1		PPB	14 Nov 2023 09:14:11	U	SMPL
CRA - 1	CRA	242	0.0454	-245.1		PPB	14 Nov 2023 09:23:29	U	SMPL
HighStd - 1	HighStd	46645	10.0629	-245.1		PPB	14 Nov 2023 09:25:44	U	SMPL
ChkStd - 1	ChkStd	33814	7.2929	-245.1		PPB	14 Nov 2023 09:27:59	U	SMPL
PB157134BL - 1	PBW	-42	-0.0159	-245.1		PPB	14 Nov 2023 09:33:04	U	SMPL
PB157134BS - 1	LCSW	19041	4.1037	-245.1		PPB	14 Nov 2023 09:35:19	U	SMPL
O5211-07 - 1	14B-(1-4)-COMP	5598	1.2016	-245.1		PPB	14 Nov 2023 09:37:36	U	SMPL
O5231-07 - 1	14B-(1-4)-COMP	5450	1.1697	-245.1		PPB	14 Nov 2023 09:39:54	U	SMPL
O5282-01 - 1	YORK-20231106	-344	-0.0811	-245.1		PPB	14 Nov 2023 09:42:13	U	SMPL
O5283-07 - 1	14B-(1-4)-COMP	2365	0.5037	-245.1		PPB	14 Nov 2023 09:44:33	U	SMPL
O5283-07DUP - 1	14B-(1-4)-COMPDUP	2318	0.4936	-245.1		PPB	14 Nov 2023 09:46:48	U	SMPL
CCV18 - 1	CCV18	22200	4.7857	-245.1		PPB	14 Nov 2023 09:49:04	U	SMPL
CCB18 - 1	CCB18	-76	-0.0233	-245.1		PPB	14 Nov 2023 09:51:20	U	SMPL
O5283-07MS - 1	14B-(1-4)-COMPMS	19364	4.1735	-245.1		PPB	14 Nov 2023 09:53:35	U	SMPL
O5283-07MSD - 1	14B-(1-4)-COMPMSD	19002	4.0953	-245.1		PPB	14 Nov 2023 09:55:51	U	SMPL
O5298-07 - 1	14B-(1-4)-COMP	2225	0.4735	-245.1		PPB	14 Nov 2023 09:58:08	U	SMPL
O5310-01 - 1	SEMI-ANNUAL-CONDENSATE	763	0.1579	-245.1		PPB	14 Nov 2023 10:00:26	U	SMPL
O5283-07LX5 - 1		449	0.0901	-245.1		PPB	14 Nov 2023 10:02:42	U	SMPL
O5283-07A - 1		18199	3.9220	-245.1		PPB	14 Nov 2023 10:04:58	U	SMPL
CCV19 - 1	CCV19	22195	4.7846	-245.1		PPB	14 Nov 2023 10:07:14	U	SMPL
CCB19 - 1	CCB19	-118	-0.0323	-245.1		PPB	14 Nov 2023 10:09:33	U	SMPL