

LB128319

7470A

INSTRUMENT ID: 1012

Linear

 μ Abs.:

45684

A= 0.0000e+000

B= 2.2269e-004

C= -1.0347e-002

Rho= 0.9994570

Accept=Accepted

Max. Conc: —

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(u)
0.0	0.000	0.053	0.053	286	0.000	286	0				14
0.2	0.200	0.227	0.027	1065	0.0 %	1065					-1
2.5	2.500	2.466	-0.034	11118	0.0 %	11118					1
5.0	5.000	5.029	0.029	22630	0.0 %	22630					-3
7.5	7.500	7.262	-0.238	32658	0.0 %	32658					2
10.0	10.000	10.163	0.163	45684	0.0 %	45684					

LB128313 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Nov 2023 11:38:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	286	-	0.0000	7470A	PPB	14 Nov 2023 12:03:27	S	Std
0.2 - 1	50.2	1065	-	0.2000	7470A	PPB	14 Nov 2023 12:05:44	S	Std
2.5 - 1	52.5	11118	-	2.5000	7470A	PPB	14 Nov 2023 12:08:01	S	Std
5.0 - 1	55	22630	-	5.0000	7470A	PPB	14 Nov 2023 12:10:18	S	Std
7.5 - 1	57.5	32658	-	7.5000	7470A	PPB	14 Nov 2023 12:17:57	S	Std
10.0 - 1	57.1	45684	-	10.0000	7470A	PPB	14 Nov 2023 12:22:46	S	Std
ICV72 - 1	ICV72	18103	4.0210	-	7470A	PPB	14 Nov 2023 12:28:27	U	SMPL
ICB72 - 1	ICB72	-271	-0.0707	-	7470A	PPB	14 Nov 2023 12:30:43	U	SMPL
CCV20 - 1	CCV20	22281	4.9514	-	7470A	PPB	14 Nov 2023 12:32:59	U	SMPL
CCB20 - 1	CCB20	-281	-0.0729	-	7470A	PPB	14 Nov 2023 12:35:14	U	SMPL
CRA - 1	CRA	919	0.1943	-	7470A	PPB	14 Nov 2023 12:37:30	U	SMPL
HighStd - 1	HighStd	44861	9.9797	-	7470A	PPB	14 Nov 2023 12:39:46	U	SMPL
ChkStd - 1	ChkStd	30006	6.6717	-	7470A	PPB	14 Nov 2023 12:42:01	U	SMPL
PB157128BL - 1	PBW	-376	-0.0941	-	7470A	PPB	14 Nov 2023 12:44:19	U	SMPL
PB157128BS - 1	LCSW	16906	3.7544	-	7470A	PPB	14 Nov 2023 12:49:39	U	SMPL
O5359-02 - 1	48-WC-1	-205	-0.0560	-	7470A	PPB	14 Nov 2023 12:51:54	U	SMPL
O5359-02DUP - 1	48-WC-1DUP	-40	-0.0193	-	7470A	PPB	14 Nov 2023 12:54:12	U	SMPL
O5359-02MS - 1	48-WC-1MS	18074	4.0145	-	7470A	PPB	14 Nov 2023 12:56:29	U	SMPL
O5359-02MSD - 1	48-WC-1MSD	18404	4.0880	-	7470A	PPB	14 Nov 2023 12:58:47	U	SMPL
O5376-02 - 1	204	-153	-0.0444	-	7470A	PPB	14 Nov 2023 13:01:03	U	SMPL
CCV21 - 1	CCV21	24309	5.4030	-	7470A	PPB	14 Nov 2023 13:23:57	U	SMPL
CCB21 - 1	CCB21	-88	-0.0299	-	7470A	PPB	14 Nov 2023 13:26:12	U	SMPL
O5376-04 - 1	210	5	-0.0092	-	7470A	PPB	14 Nov 2023 13:28:28	U	SMPL
O5376-06 - 1	223	20	-0.0059	-	7470A	PPB	14 Nov 2023 13:30:44	U	SMPL
O5378-04 - 1	MH-3	21	-0.0057	-	7470A	PPB	14 Nov 2023 13:33:00	U	SMPL
O5379-04 - 1	WC-9	57	0.0023	-	7470A	PPB	14 Nov 2023 13:35:16	U	SMPL
O5379-08 - 1	WC-4	45	-0.0003	-	7470A	PPB	14 Nov 2023 13:37:31	U	SMPL
O5379-12 - 1	WC-5	50	0.0008	-	7470A	PPB	14 Nov 2023 13:39:47	U	SMPL
O5380-04 - 1	WC-7	70	0.0052	-	7470A	PPB	14 Nov 2023 13:42:04	U	SMPL
O5380-08 - 1	WC-8	48	0.0003	-	7470A	PPB	14 Nov 2023 13:44:21	U	SMPL
O5386-02 - 1	CONTAMINATED-PLASTIC	132	0.0190	-	7470A	PPB	14 Nov 2023 13:46:38	U	SMPL
PB157129BL - 1	PBW	31	-0.0034	-	7470A	PPB	14 Nov 2023 13:48:54	U	SMPL
CCV22 - 1	CCV22	23225	5.1616	-	7470A	PPB	14 Nov 2023 13:51:12	U	SMPL
CCB22 - 1	CCB22	-209	-0.0569	-	7470A	PPB	14 Nov 2023 13:53:28	U	SMPL
PB157129BS - 1	LCSW	17813	3.9564	-	7470A	PPB	14 Nov 2023 13:55:46	U	SMPL
O5386-01 - 1	CONTAMINATED-WATER	7775	1.7211	-	7470A	PPB	14 Nov 2023 14:03:12	U	SMPL
O5386-01DUP - 1	CONTAMINATED-WATERDUP	7618	1.6861	-	7470A	PPB	14 Nov 2023 14:05:27	U	SMPL
O5386-01MS - 1	CONTAMINATED-WATERMS	21960	4.8799	-	7470A	PPB	14 Nov 2023 14:07:45	U	SMPL
O5386-01MSD - 1	CONTAMINATED-WATERMSD	20825	4.6272	-	7470A	PPB	14 Nov 2023 14:10:02	U	SMPL
PB157074TB - 1	PB157074TB	67	0.0046	-	7470A	PPB	14 Nov 2023 14:15:00	U	SMPL
PB157113TB - 1	PB157113TB	67	0.0046	-	7470A	PPB	14 Nov 2023 14:17:17	U	SMPL
CCV23 - 1	CCV23	20765	4.6138	-	7470A	PPB	14 Nov 2023 14:22:36	U	SMPL
CCB23 - 1	CCB23	-47	-0.0208	-	7470A	PPB	14 Nov 2023 14:24:53	U	SMPL
O5359-02LX5 - 1		-29	-0.0168	-	7470A	PPB	14 Nov 2023 14:27:10	U	SMPL
O5359-02A - 1		17828	3.9598	-	7470A	PPB	14 Nov 2023 14:29:26	U	SMPL
O5386-01LX5 - 1		1713	0.3711	-	7470A	PPB	14 Nov 2023 14:31:41	U	SMPL
O5386-01A - 1		22700	5.0447	-	7470A	PPB	14 Nov 2023 14:33:58	U	SMPL
CCV24 - 1	CCV24	20640	4.5860	-	7470A	PPB	14 Nov 2023 14:39:36	U	SMPL
CCB24 - 1	CCB24	-322	-0.0821	-	7470A	PPB	14 Nov 2023 14:41:51	U	SMPL