



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128315

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	11/14/2023	12:10
2	O5386-01		1	12	100.00	>212.0	>212.0	11/14/2023	12:40
3	O5386-01DUP		1	12	100.00	>212.0	>212.0	11/14/2023	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB128315

WorkList Name : fp-11-13

WorkList ID : 175499

Department : Wet-Chemistry

Date : 11-13-2023 08:08:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5386-01	CONTAMINATED-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	L41	11/10/2023	1010B

Date/Time 11/14/2023 12:00
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: 701 Gdc

Date/Time 11/14/2023 13:00
 Raw Sample Received by: 701 Gdc
 Raw Sample Relinquished by: 701 Gdc