



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128319

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	11/14/2023	10:20
2	CAL2	1	Water	NA	NA	20.4	7.00	11/14/2023	10:21
3	CAL3	1	Water	NA	NA	20.4	10.02	11/14/2023	10:25
4	ICV	1	Water	NA	NA	20.7	7.01	11/14/2023	10:30
5	CCV1	1	Water	NA	NA	20.6	2.01	11/14/2023	10:33
6	O5362-01	1	Water	NA	NA	20.7	5.39	11/14/2023	10:38
7	O5362-01DUP	1	Water	NA	NA	20.8	5.40	11/14/2023	10:39
8	CCV2	1	Water	NA	NA	20.7	12.02	11/14/2023	10:40

✓ 128319

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O5362 WorkList ID : 175555 Department : Wet-Chemistry Date : 11-14-2023 10:05:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5362-01	A POOL-WATER	Water	pH	Cool 4 deg C	STER05	L41	11/09/2023	9040C

Date/Time 11-14-23 10:10
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 11-14-23 11:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]