



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128331

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	11/15/2023	13:30
2	CAL2	1	Water	NA	NA	20.6	7.00	11/15/2023	13:31
3	CAL3	1	Water	NA	NA	20.5	10.02	11/15/2023	13:32
4	ICV	1	Water	NA	NA	20.5	7.01	11/15/2023	13:35
5	CCV1	1	Water	NA	NA	20.7	2.01	11/15/2023	13:40
6	O5417-01	1	Water	NA	NA	22.5	7.45	11/15/2023	13:45
7	O5435-01	1	Water	NA	NA	20.1	4.90	11/15/2023	13:55
8	O5435-01DUP	1	Water	NA	NA	20.2	4.91	11/15/2023	13:56
9	CCV2	1	Water	NA	NA	20.2	12.02	11/15/2023	14:00

12833)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O5435 WorkList ID : 175601 Department : Wet-Chemistry Date : 11-15-2023 13:01:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5417-01	1113	Water	pH	Cool 4 deg C	PSEG03	L41	11/14/2023	9040C
O5435-01	A-B-CELL-POOL-WATER-3RD-	Water	pH	Cool 4 deg C	STER10	L41	11/14/2023	9040C

Date/Time 11-15-23 13:20
Raw Sample Received by: 36/0201
Raw Sample Relinquished by: 12/01

Date/Time 11-15-23 15:00
Raw Sample Received by: 12/01
Raw Sample Relinquished by: 36/0201