


LB12833T

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.7821e-004 *slope*C= -1.0444e-001 *y-intercept*

Rho= 0.9998202

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/d</i>
0.00	0.000	0.096	0.096	1124	0.000	1124					1
0.05	0.050					419					1
0.20	0.200	0.157	-0.043	1469	0.0 %	1469	0				-22
2.50	2.500	2.515	0.015	14700	0.0 %	14700					1
5.00	5.000	4.884	-0.116	27992	0.0 %	27992					-2
7.50	7.500	7.481	-0.019	42566	0.0 %	42566					0
10.0	10.000	10.066	0.066	57069	0.0 %	57069					1

LB1283345

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	1124	-		0 SFAM01.1	PPB	11/15/2023 14:49	Std
	0.2 S01	1469	-		0.2 SFAM01.1	PPB	11/15/2023 14:51	Std
	2.5 S02	14700	-		2.5 SFAM01.1	PPB	11/15/2023 14:53	Std
	5 S03	27992	-		5 SFAM01.1	PPB	11/15/2023 14:56	Std
	7.5 S04	42566	-		7.5 SFAM01.1	PPB	11/15/2023 14:58	Std
	10 S05	57069	-		10 SFAM01.1	PPB	11/15/2023 15:09	Std
ICV074	ICV074	22345	3.8777 -		SFAM01.1	PPB	11/15/2023 15:11	SMPL
ICB074	ICB074	783	0.0351 -		SFAM01.1	PPB	11/15/2023 15:14	SMPL
CCV029	CCV029	28245	4.9292 -		SFAM01.1	PPB	11/15/2023 15:16	SMPL
CCB029	CCB029	778	0.0342 -		SFAM01.1	PPB	11/15/2023 15:18	SMPL
PB157185BL	PBW185	117	-0.0836 -		SFAM01.1	PPB	11/15/2023 15:21	SMPL
O5058-01	MA49G7	109	-0.085 -		SFAM01.1	PPB	11/15/2023 15:23	SMPL
O5058-04	MA4A81	28856	5.0381 -		SFAM01.1	PPB	11/15/2023 15:25	SMPL
O5058-05	MA49F7	-124	-0.1265 -		SFAM01.1	PPB	11/15/2023 15:27	SMPL
O5058-06	MA4972	161	-0.0758 -		SFAM01.1	PPB	11/15/2023 15:30	SMPL
O5058-07	MA49W9	185	-0.0715 -		SFAM01.1	PPB	11/15/2023 15:32	SMPL
O5058-08	MA49W5	184	-0.0717 -		SFAM01.1	PPB	11/15/2023 15:34	SMPL
O5058-09	MA49Z5	134	-0.0806 -		SFAM01.1	PPB	11/15/2023 15:36	SMPL
O5058-10	MA49Y5	323	-0.0469 -		SFAM01.1	PPB	11/15/2023 15:39	SMPL
O5058-11	MA49Y8	213	-0.0665 -		SFAM01.1	PPB	11/15/2023 15:41	SMPL
O5058-12	MA49Y7	192	-0.0702 -		SFAM01.1	PPB	11/15/2023 15:43	SMPL
O5058-13	MA49Y7D	171	-0.074 -		SFAM01.1	PPB	11/15/2023 15:45	SMPL
O5058-14	MA49Y7S	5176	0.818 -		SFAM01.1	PPB	11/15/2023 15:48	SMPL
PB157186BL	PBW186	236	-0.0624 -		SFAM01.1	PPB	11/15/2023 15:50	SMPL
O5059-01	MA49F6	315	-0.0483 -		SFAM01.1	PPB	11/15/2023 15:52	SMPL
O5059-03	MA4A68	20745	3.5926 -		SFAM01.1	PPB	11/15/2023 15:54	SMPL
O5059-04	MA49G8	150	-0.0777 -		SFAM01.1	PPB	11/15/2023 15:57	SMPL
O5059-05	MA4973	140	-0.0795 -		SFAM01.1	PPB	11/15/2023 15:59	SMPL
O5059-06	MA49W4	347	-0.0426 -		SFAM01.1	PPB	11/15/2023 16:01	SMPL
O5059-07	MA49X0	256	-0.0588 -		SFAM01.1	PPB	11/15/2023 16:04	SMPL
CCV030	CCV030	27144	4.733 -		SFAM01.1	PPB	11/15/2023 16:06	SMPL
CCB030	CCB030	743	0.028 -		SFAM01.1	PPB	11/15/2023 16:08	SMPL
O5059-08	MA49Y9	169	-0.0743 -		SFAM01.1	PPB	11/15/2023 16:10	SMPL
O5059-09	MA49Y6	247	-0.0604 -		SFAM01.1	PPB	11/15/2023 16:13	SMPL
O5059-10	MA49Z4	370	-0.0385 -		SFAM01.1	PPB	11/15/2023 16:15	SMPL
O5059-11	MA49Z3	520	-0.0118 -		SFAM01.1	PPB	11/15/2023 16:17	SMPL
O5059-12	MA49Z3D	1049	0.0825 -		SFAM01.1	PPB	11/15/2023 16:19	SMPL
O5059-13	MA49Z3S	5977	0.9607 -		SFAM01.1	PPB	11/15/2023 16:22	SMPL
CCV031	CCV031	27609	4.8158 -		SFAM01.1	PPB	11/15/2023 16:38	SMPL
CCB031	CCB031	715	0.023 -		SFAM01.1	PPB	11/15/2023 16:41	SMPL