

7470A


L3128349

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55585



A= 0.0000e+000

B= 1.8066e-004

C= 2.6411e-002

Rho= 0.9999204

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Grad
0.0	0.000	0.048	0.048	120	0.000	120					
0.2	0.200	0.208	0.008	1005	0.0 %	1005	0				
2.5	2.500	2.482	-0.018	13591	0.0 %	13591					4
5.0	5.000	4.942	-0.058	27212	0.0 %	27212					-1
7.5	7.500	7.452	-0.048	41102	0.0 %	41102					-1
10.0	10.000	10.068	0.068	55585	0.0 %	55585					-1

LB128349 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Nov 2023 14:43:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	120	-	0.0000	7470A	PPB	16 Nov 2023 14:50:27	S	Std
0.2 - 1	50.2	1005	-	0.2000	7470A	PPB	16 Nov 2023 14:52:45	S	Std
2.5 - 1	52.7	13591	-	2.5000	7470A	PPB	16 Nov 2023 14:55:03	S	Std
5.0 - 1	55	27212	-	5.0000	7470A	PPB	16 Nov 2023 14:57:19	S	Std
7.5 - 1	57.5	41102	-	7.5000	7470A	PPB	16 Nov 2023 14:59:36	S	Std
10.0 - 1	57	55585	-	10.0000	7470A	PPB	16 Nov 2023 15:01:53	S	Std
ICV76 - 1	ICV76	22188	4.0348	-	7470A	PPB	16 Nov 2023 15:04:41	U	SMPL
ICB76 - 1	ICB76	-259	-0.0204	-	7470A	PPB	16 Nov 2023 15:06:57	U	SMPL
CCV36 - 1	CCV36	27218	4.9435	-	7470A	PPB	16 Nov 2023 15:35:04	U	SMPL
CCB36 - 1	CCB36	-236	-0.0162	-	7470A	PPB	16 Nov 2023 15:37:19	U	SMPL
CRI - 1	CRI	909	0.1906	-	7470A	PPB	16 Nov 2023 15:39:35	U	SMPL
HIGH STD - 1	HIGH STD	54813	9.9287	-	7470A	PPB	16 Nov 2023 15:41:49	U	SMPL
CHK STD - 1	CHK STD	37657	6.8293	-	7470A	PPB	16 Nov 2023 15:44:05	U	SMPL
PB157209BL - 1	PBW	-388	-0.0437	-	7470A	PPB	16 Nov 2023 15:46:23	U	SMPL
PB157209BS - 1	LCSW	23010	4.1833	-	7470A	PPB	16 Nov 2023 15:48:40	U	SMPL
O5415-04 - 1	RD-SW-01-20231103	548	0.1254	-	7470A	PPB	16 Nov 2023 15:50:56	U	SMPL
O5415-04DUP - 1	RD-SW-01-20231103DUP	655	0.1447	-	7470A	PPB	16 Nov 2023 15:53:14	U	SMPL
O5415-04MS - 1	RD-SW-01-20231103MS	24866	4.5186	-	7470A	PPB	16 Nov 2023 15:55:31	U	SMPL
O5415-04MSD - 1	RD-SW-01-20231103MSD	21302	3.8747	-	7470A	PPB	16 Nov 2023 15:57:48	U	SMPL
CCV37 - 1	CCV37	27728	5.0356	-	7470A	PPB	16 Nov 2023 16:02:39	U	SMPL
CCB37 - 1	CCB37	-318	-0.0310	-	7470A	PPB	16 Nov 2023 16:04:54	U	SMPL
O5415-04LX5 - 1		220	0.0662	-	7470A	PPB	16 Nov 2023 16:07:11	U	SMPL
O5415-04A - 1		23387	4.2514	-	7470A	PPB	16 Nov 2023 16:09:26	U	SMPL
CCV38 - 1	CCV38	32337	5.8683	-	7470A	PPB	16 Nov 2023 16:11:41	U	SMPL
CCB38 - 1	CCB38	-337	-0.0345	-	7470A	PPB	16 Nov 2023 16:13:58	U	SMPL