



Analytical Summary Report

Analysis Method: 9045D

Parameter: Corrosivity

Run Number: LB128351

BalanceID: WC SC-4

Analyst By : JIGNESH

Supervisor Review By : Sohil

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/16/2023	15:05
2	CAL2	1	Water	NA	NA	20.3	7.00	11/16/2023	15:06
3	CAL3	1	Water	NA	NA	20.3	10.02	11/16/2023	15:10
4	ICV	1	Water	NA	NA	20.6	7.02	11/16/2023	15:15
5	CCV1	1	Water	NA	NA	20.6	2.01	11/16/2023	15:20
6	O5445-03	1	Solid	20.02	20	20.9	6.76	11/16/2023	15:25
7	O5445-03DUP	1	Solid	20.03	20	20.8	6.77	11/16/2023	15:26
8	O5445-06	1	Solid	20.04	20	24.4	7.51	11/16/2023	15:35
9	O5445-09	1	Solid	20.02	20	24.3	7.64	11/16/2023	15:39
10	O5447-03	1	Solid	20.03	20	24.4	7.80	11/16/2023	15:40
11	O5447-06	1	Solid	20.02	20	21.8	7.38	11/16/2023	15:45
12	O5449-03	1	Solid	20.02	20	21.2	10.99	11/16/2023	15:50
13	O5449-06	1	Solid	20.03	20	21.8	7.39	11/16/2023	16:00
14	O5449-09	1	Solid	20.04	20	22.2	7.23	11/16/2023	16:10
15	O5449-12	1	Solid	20.03	20	21.1	12.05	11/16/2023	16:15
16	CCV2	1	Water	NA	NA	20.7	12.02	11/16/2023	16:20
17	O5451-03	1	Solid	20.04	20	22.3	8.50	11/16/2023	16:25
18	O5466-03	1	Solid	20.04	20	25.6	8.62	11/16/2023	16:30
19	O5466-06	1	Solid	20.02	20	24.8	8.10	11/16/2023	16:35
20	CCV3	1	Water	NA	NA	20.8	2.01	11/16/2023	16:40