


LB128358

7471B

INSTRUMENT ID: CV1

μ Abs.:

58308



Linear

A= 0.0000e+000

B= 1.7028e-004

C= -3.3049e-003

Rho= 0.9998211

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%)
0.0	0.000	0.016	0.016	114	0.000	114					
0.2	0.200	0.188	-0.012	1124	0.0 %	1124					-6
2.5	2.500	2.492	-0.008	14656	0.0 %	14656					0
5.0	5.000	4.939	-0.061	29024	0.0 %	29024					-1
7.5	7.500	7.639	0.139	44884	0.0 %	44884					2
10.0	10.000	9.925	-0.075	58308	0.0 %	58308					-1

LB128358 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Nov 2023 07:44:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	114	-	0.0000	7471B	PPB	17 Nov 2023 07:53:02	S	Std
0.2 - 1	50.2	1124	-	0.2000	7471B	PPB	17 Nov 2023 07:55:19	S	Std
2.5 - 1	52.5	14656	-	2.5000	7471B	PPB	17 Nov 2023 07:57:36	S	Std
5.0 - 1	55	29024	-	5.0000	7471B	PPB	17 Nov 2023 07:59:53	S	Std
7.5 - 1	57.5	44884	-	7.5000	7471B	PPB	17 Nov 2023 08:02:09	S	Std
10.0 - 1	510	58308	-	10.0000	7471B	PPB	17 Nov 2023 08:07:02	S	Std
ICV77 - 1	ICV77	24815	4.2221	-	7471B	PPB	17 Nov 2023 08:09:49	U	SMPL
ICB77 - 1	ICB77	-218	-0.0404	-	7471B	PPB	17 Nov 2023 08:12:05	U	SMPL
CCV39 - 1	CCV39	28664	4.8775	-	7471B	PPB	17 Nov 2023 08:14:22	U	SMPL
CCB39 - 1	CCB39	-200	-0.0374	-	7471B	PPB	17 Nov 2023 08:16:37	U	SMPL
HighStd - 1	HighStd	58666	9.9862	-	7471B	PPB	17 Nov 2023 08:21:08	U	SMPL
ChkStd - 1	ChkStd	42776	7.2805	-	7471B	PPB	17 Nov 2023 08:23:24	U	SMPL
CRA - 1	CRA	1230	0.2061	-	7471B	PPB	17 Nov 2023 08:26:11	U	SMPL
PB157212BL - 1	PBS	-156	-0.0299	-	7471B	PPB	17 Nov 2023 08:28:41	U	SMPL
PB157212BS - 1	LCSS	22260	3.7871	-	7471B	PPB	17 Nov 2023 08:30:56	U	SMPL
O5294-05 - 1	WT-5	5517	0.9361	-	7471B	PPB	17 Nov 2023 08:33:13	U	SMPL
O5418-01 - 1	STONE-1	713	0.1181	-	7471B	PPB	17 Nov 2023 08:35:31	U	SMPL
O5418-04 - 1	STONE-2	280	0.0444	-	7471B	PPB	17 Nov 2023 08:37:51	U	SMPL
O5418-07 - 1	STONE-3	49	0.0050	-	7471B	PPB	17 Nov 2023 08:40:08	U	SMPL
O5420-01 - 1	72-12018	4333	0.7345	-	7471B	PPB	17 Nov 2023 08:42:23	U	SMPL
CCV40 - 1	CCV40	29731	5.0592	-	7471B	PPB	17 Nov 2023 08:44:38	U	SMPL
CCB40 - 1	CCB40	-130	-0.0254	-	7471B	PPB	17 Nov 2023 08:46:55	U	SMPL
O5425-01 - 1	O-1(20)(15-20)	1575	0.2649	-	7471B	PPB	17 Nov 2023 08:49:11	U	SMPL
O5425-02 - 1	O-5(5-10)	801	0.1331	-	7471B	PPB	17 Nov 2023 08:56:22	U	SMPL
O5425-02DUP - 1	O-5(5-10)DUP	704	0.1166	-	7471B	PPB	17 Nov 2023 08:58:37	U	SMPL
O5425-03 - 1	O-6(15-20)	478	0.0781	-	7471B	PPB	17 Nov 2023 09:00:53	U	SMPL
O5425-09 - 1	O5425-02MS	26896	4.5765	-	7471B	PPB	17 Nov 2023 09:03:09	U	SMPL
O5425-10 - 1	O5425-02MSD	24928	4.2414	-	7471B	PPB	17 Nov 2023 09:05:25	U	SMPL
O5432-01 - 1	TP-6	4717	0.7999	-	7471B	PPB	17 Nov 2023 09:07:43	U	SMPL
O5432-05 - 1	TP-7	7398	1.2564	-	7471B	PPB	17 Nov 2023 09:10:01	U	SMPL
O5432-09 - 1	TP-8	5888	0.9993	-	7471B	PPB	17 Nov 2023 09:12:19	U	SMPL
O5445-01 - 1	SOIL-1-1115	5759	0.9773	-	7471B	PPB	17 Nov 2023 09:14:34	U	SMPL
CCV41 - 1	CCV41	30133	5.1277	-	7471B	PPB	17 Nov 2023 09:16:50	U	SMPL
CCB41 - 1	CCB41	-32	-0.0088	-	7471B	PPB	17 Nov 2023 09:19:05	U	SMPL
O5445-04 - 1	SOIL-2-1115	132873	22.6220	-	7471B	PPB	17 Nov 2023 09:21:23	U	SMPL
O5445-07 - 1	SOIL-3-1115	32550	5.5392	-	7471B	PPB	17 Nov 2023 09:23:38	U	SMPL
O5446-01 - 1	JC-03-111523A	3676	0.6226	-	7471B	PPB	17 Nov 2023 09:25:59	U	SMPL
O5446-03 - 1	JC-03-111523B	46009	7.8310	-	7471B	PPB	17 Nov 2023 09:28:16	U	SMPL
O5447-01 - 1	SP1	1400	0.2351	-	7471B	PPB	17 Nov 2023 09:30:32	U	SMPL
O5447-04 - 1	SP2	1047	0.1750	-	7471B	PPB	17 Nov 2023 09:32:50	U	SMPL
O5448-01 - 1	OR-02-111523	2235	0.3773	-	7471B	PPB	17 Nov 2023 09:35:07	U	SMPL
O5425-02LX5 - 1		88	0.0117	-	7471B	PPB	17 Nov 2023 09:37:24	U	SMPL
O5425-02A - 1		25651	4.3645	-	7471B	PPB	17 Nov 2023 09:49:05	U	SMPL
O5445-04DLX5 - 1		29825	5.0752	-	7471B	PPB	17 Nov 2023 09:51:21	U	SMPL
CCV42 - 1	CCV42	30190	5.1374	-	7471B	PPB	17 Nov 2023 09:53:37	U	SMPL
CCB42 - 1	CCB42	-50	-0.0118	-	7471B	PPB	17 Nov 2023 09:55:53	U	SMPL