

LB128360


7470A

INSTRUMENT 1A: CV1

Linear

μ Abs.:

53092



A= 0.0000e+000

B= 1.8768e-004

C= 6.3083e-002

Rho= 0.9996993

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%)
0.0	0.000	0.077	0.077	75	0.000	75	0				
0.2	0.200	0.227	0.027	871	0.0 %	871					14
2.5	2.500	2.458	-0.042	12763	0.0 %	12763					-2
5.0	5.000	4.821	-0.179	25352	0.0 %	25352					-4
7.5	7.500	7.589	0.089	40102	0.0 %	40102					1
10.0	10.000	10.027	0.027	53092	0.0 %	53092					0

LB128360 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Nov 2023 10:38:06

Sample ID	Extended ID	μ Abs.	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	75	-	0.0000	7470A	PPB	17 Nov 2023 10:38:16	S	Std
0.2 - 1	50.2	871	-	0.2000	7470A	PPB	17 Nov 2023 10:40:33	S	Std
2.5 - 1	52.5	12763	-	2.5000	7470A	PPB	17 Nov 2023 10:42:50	S	Std
5.0 - 1	55	25352	-	5.0000	7470A	PPB	17 Nov 2023 10:45:07	S	Std
7.5 - 1	57.5	40102	-	7.5000	7470A	PPB	17 Nov 2023 10:47:24	S	Std
10.0 - 1	510	53092	-	10.0000	7470A	PPB	17 Nov 2023 10:49:39	S	Std
ICV78 - 1	ICV78	21596	4.1162	-7470A	PPB	17 Nov 2023 10:52:36	U	SMPL	
ICB78 - 1	ICB78	-217	0.0224	-7470A	PPB	17 Nov 2023 10:54:51	U	SMPL	
CCV43 - 1	CCV43	25338	4.8185	-7470A	PPB	17 Nov 2023 10:57:06	U	SMPL	
CCB43 - 1	CCB43	-219	0.0220	-7470A	PPB	17 Nov 2023 10:59:21	U	SMPL	
CRA - 1	CRA	819	0.2168	-7470A	PPB	17 Nov 2023 11:01:38	U	SMPL	
HighStd - 1	HighStd	51973	9.8173	-7470A	PPB	17 Nov 2023 11:03:53	U	SMPL	
ChkStd - 1	ChkStd	37945	7.1846	-7470A	PPB	17 Nov 2023 11:06:08	U	SMPL	
PB157188BL - 1	PBW	-424	-0.0165	-7470A	PPB	17 Nov 2023 11:08:27	U	SMPL	
PB157188BS - 1	LCSW	22527	4.2909	-7470A	PPB	17 Nov 2023 11:10:44	U	SMPL	
O5412-04 - 1	MH-1	-315	0.0040	-7470A	PPB	17 Nov 2023 11:13:01	U	SMPL	
O5412-08 - 1	MH-2	-106	0.0432	-7470A	PPB	17 Nov 2023 11:15:18	U	SMPL	
O5417-03 - 1	1114	27	0.0681	-7470A	PPB	17 Nov 2023 11:17:35	U	SMPL	
O5417-04 - 1	1115	54	0.0732	-7470A	PPB	17 Nov 2023 11:19:52	U	SMPL	
O5417-05 - 1	1116	533	0.1631	-7470A	PPB	17 Nov 2023 11:22:07	U	SMPL	
CCV44 - 1	CCV44	24193	4.6036	-7470A	PPB	17 Nov 2023 11:24:23	U	SMPL	
CCB44 - 1	CCB44	-248	0.0165	-7470A	PPB	17 Nov 2023 11:26:38	U	SMPL	
O5417-06 - 1	1117	24	0.0676	-7470A	PPB	17 Nov 2023 11:28:55	U	SMPL	
O5417-07 - 1	1118	34	0.0695	-7470A	PPB	17 Nov 2023 11:31:10	U	SMPL	
O5418-03 - 1	STONE-1	-27	0.0580	-7470A	PPB	17 Nov 2023 11:33:26	U	SMPL	
O5418-06 - 1	STONE-2	14	0.0657	-7470A	PPB	17 Nov 2023 11:35:42	U	SMPL	
O5418-09 - 1	STONE-3	22	0.0672	-7470A	PPB	17 Nov 2023 11:37:58	U	SMPL	
O5432-04 - 1	TP-6	17	0.0663	-7470A	PPB	17 Nov 2023 11:40:14	U	SMPL	
O5432-08 - 1	TP-7	18	0.0665	-7470A	PPB	17 Nov 2023 11:42:31	U	SMPL	
O5432-12 - 1	TP-8	57	0.0738	-7470A	PPB	17 Nov 2023 11:44:47	U	SMPL	
O5432-12DUP - 1	TP-8DUP	52	0.0728	-7470A	PPB	17 Nov 2023 11:47:05	U	SMPL	
O5432-12MS - 1	TP-8MS	20434	3.8981	-7470A	PPB	17 Nov 2023 11:49:21	U	SMPL	
CCV45 - 1	CCV45	24208	4.6064	-7470A	PPB	17 Nov 2023 11:51:36	U	SMPL	
CCB45 - 1	CCB45	-247	0.0167	-7470A	PPB	17 Nov 2023 11:53:53	U	SMPL	
O5432-12MSD - 1	TP-8MSD	21874	4.1684	-7470A	PPB	17 Nov 2023 11:56:11	U	SMPL	
PB157209BL - 1	PBW	-249	0.0164	-7470A	PPB	17 Nov 2023 11:58:26	U	SMPL	
PB157209BS - 1	LCSW	21922	4.1774	-7470A	PPB	17 Nov 2023 12:00:43	U	SMPL	
O5415-04 - 1	RD-SW-01-20231103	565	0.1691	-7470A	PPB	17 Nov 2023 12:02:59	U	SMPL	
O5415-04DUP - 1	RD-SW-01-20231103DUP	638	0.1828	-7470A	PPB	17 Nov 2023 12:08:04	U	SMPL	
O5415-04MS - 1	RD-SW-01-20231103MS	24321	4.6276	-7470A	PPB	17 Nov 2023 12:10:19	U	SMPL	
O5415-04MSD - 1	RD-SW-01-20231103MSD	20671	3.9426	-7470A	PPB	17 Nov 2023 12:12:36	U	SMPL	
O5417-01 - 1	1113	5871	1.1649	-7470A	PPB	17 Nov 2023 12:14:54	U	SMPL	
CCV46 - 1	CCV46	25993	4.9414	-7470A	PPB	17 Nov 2023 12:17:12	U	SMPL	
CCB46 - 1	CCB46	-179	0.0295	-7470A	PPB	17 Nov 2023 12:19:31	U	SMPL	
O5432-12LX5 - 1		-49	0.0539	-7470A	PPB	17 Nov 2023 12:21:47	U	SMPL	
O5432-12A - 1		19956	3.8084	-7470A	PPB	17 Nov 2023 12:24:03	U	SMPL	
O5415-04LX5 - 1		11	0.0651	-7470A	PPB	17 Nov 2023 12:26:18	U	SMPL	
O5415-04A - 1		21152	4.0329	-7470A	PPB	17 Nov 2023 12:28:34	U	SMPL	
PB157152TB - 1	PB157152TB	-206	0.0244	-7470A	PPB	17 Nov 2023 12:30:49	U	SMPL	
CCV47 - 1	CCV47	25835	4.9118	-7470A	PPB	17 Nov 2023 12:33:05	U	SMPL	
CCB47 - 1	CCB47	-210	0.0237	-7470A	PPB	17 Nov 2023 12:35:22	U	SMPL	