



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128361

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	11/17/2023	08:50
2	O5417-01		1	14	100.00	>212.0	>212.0	11/17/2023	09:20
3	O5417-01DUP		1	14	100.00	>212.0	>212.0	11/17/2023	09:50
4	O5460-01		1	12	100.00	>212.0	>212.0	11/17/2023	10:20
5	O5460-02		1	14	100.00	>212.0	>212.0	11/17/2023	10:50
6	O5460-03		1	12	100.00	>212.0	>212.0	11/17/2023	11:20
7	O5460-04		1	12	100.00	>212.0	>212.0	11/17/2023	11:50
8	O5460-05		1	14	100.00	>212.0	>212.0	11/17/2023	12:20
9	O5460-06		1	13	100.00	>212.0	>212.0	11/17/2023	12:50
10	O5460-07		1	14	100.00	>212.0	>212.0	11/17/2023	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB128361

WorkList Name : fp11-15

WorkList ID : 175590

Department : Wet-Chemistry

Date : 11-15-2023 08:15:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5460-01D	W-A-2-QUEENS-PLAZA	Water	Flash Point	Cool 4 deg C	TULL02	L11	11/16/2023	1010B
O5460-02D	W-A-8-50TH-AVE-11TH-ST	Water	Flash Point	Cool 4 deg C	TULL02	L11	11/16/2023	1010B
O5460-03D	W-A-3-JACKSON	Water	Flash Point	Cool 4 deg C	TULL02	L11	11/16/2023	1010B
O5460-04D	W-A-3-HUNTER	Water	Flash Point	Cool 4 deg C	TULL02	L11	11/16/2023	1010B
O5460-05E	W-A-4-HUNTER	Water	Flash Point	Cool 4 deg C	TULL02	L11	11/16/2023	1010B
O5460-06E	W-A-3-CORONA	Water	Flash Point	Cool 4 deg C	TULL02	L11	11/16/2023	1010B
O5460-07D	W-A-2-CORONA	Water	Flash Point	Cool 4 deg C	TULL02	L11	11/16/2023	1010B
O5417-01E	1113	Water	Flash Point	Cool 4 deg C	PSEG03	L41	11/14/2023	1010B

Date/Time 11/17/2023 08:15
 Raw Sample Received by: RMA
 Raw Sample Relinquished by: JLC

Date/Time 11/17/2023 13:45
 Raw Sample Received by: JLC
 Raw Sample Relinquished by: RMA