


LB128369

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.8211e-004 *stop*C= 2.7315e-002 *+1.1% error*

Rho= 0.9998661

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>910D</i>
0.00	0.000	0.049	0.049	120	0.000	120					=
0.05	0.050					419	0				7
0.20	0.200	0.214	0.014	1026	0.0 %	1026					~2
2.50	2.500	2.449	-0.051	13299	0.0 %	13299					-2
5.00	5.000	4.909	-0.091	26806	0.0 %	26806					1
7.50	7.500	7.591	0.091	41532	0.0 %	41532					0
10.0	10.000	9.988	-0.012	54693	0.0 %	54693					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	120	-		0 SFAM01.1	PPB	11/17/2023 15:34	Std
	0.2 S01	1026	-		0.2 SFAM01.1	PPB	11/17/2023 15:36	Std
	2.5 S02	13299	-		2.5 SFAM01.1	PPB	11/17/2023 15:38	Std
	5 S03	26806	-		5 SFAM01.1	PPB	11/17/2023 15:42	Std
	7.5 S04	41532	-		7.5 SFAM01.1	PPB	11/17/2023 15:45	Std
	10 S05	54693	-		10 SFAM01.1	PPB	11/17/2023 15:47	Std
ICV079	ICV079	22253	4.0799 -		SFAM01.1	PPB	11/17/2023 15:50	SMPL
ICB079	ICB079	-174	-0.0044 -		SFAM01.1	PPB	11/17/2023 15:52	SMPL
CCV048	CCV048	26883	4.923 -		SFAM01.1	PPB	11/17/2023 15:54	SMPL
CCB048	CCB048	-190	-0.0073 -		SFAM01.1	PPB	11/17/2023 15:56	SMPL
PB157252BL	PBW252	-33	0.0213 -		SFAM01.1	PPB	11/17/2023 15:59	SMPL
O5115-01	MA4AH7	5195	0.9734 -		SFAM01.1	PPB	11/17/2023 16:01	SMPL
O5115-02	MA4AJ5	-1	0.0271 -		SFAM01.1	PPB	11/17/2023 16:03	SMPL
O5115-03	MA4AJ7	-27	0.0224 -		SFAM01.1	PPB	11/17/2023 16:06	SMPL
O5115-06	MA4AZ5	2791	0.5356 -		SFAM01.1	PPB	11/17/2023 16:10	SMPL
O5115-07	MA4AK1	137	0.0523 -		SFAM01.1	PPB	11/17/2023 16:12	SMPL
O5115-08	MA4AK1D	234	0.0699 -		SFAM01.1	PPB	11/17/2023 16:15	SMPL
O5115-09	MA4AK1S	5417	1.0138 -		SFAM01.1	PPB	11/17/2023 16:17	SMPL
O5115-10	MA4AK3	172	0.0586 -		SFAM01.1	PPB	11/17/2023 16:19	SMPL
O5115-11	MA4AK7	-28	0.0222 -		SFAM01.1	PPB	11/17/2023 16:21	SMPL
PB157253BL	PBW253	53	0.037 -		SFAM01.1	PPB	11/17/2023 16:24	SMPL
O5116-01	MA4AH8	510	0.1202 -		SFAM01.1	PPB	11/17/2023 16:26	SMPL
O5116-02	MA4AJ6	-32	0.0215 -		SFAM01.1	PPB	11/17/2023 16:28	SMPL
O5116-03	MA4AJ8	33	0.0333 -		SFAM01.1	PPB	11/17/2023 16:31	SMPL
O5116-06	MA4AZ4	617	0.1397 -		SFAM01.1	PPB	11/17/2023 16:35	SMPL
O5116-07	MA4AK2	-15	0.0246 -		SFAM01.1	PPB	11/17/2023 16:37	SMPL
O5116-08	MA4AK2D	30	0.0328 -		SFAM01.1	PPB	11/17/2023 16:40	SMPL
O5116-09	MA4AK2S	6127	1.1431 -		SFAM01.1	PPB	11/17/2023 16:42	SMPL
CCV049	CCV049	25856	4.736 -		SFAM01.1	PPB	11/17/2023 16:44	SMPL
CCB049	CCB049	-108	0.0076 -		SFAM01.1	PPB	11/17/2023 16:46	SMPL
O5116-10	MA4AK4	157	0.0559 -		SFAM01.1	PPB	11/17/2023 16:49	SMPL
O5116-11	MA4AZ3	-52	0.0178 -		SFAM01.1	PPB	11/17/2023 16:51	SMPL
O5115-05	MA4AL0	22231	4.0759 -		SFAM01.1	PPB	11/17/2023 16:53	SMPL
O5116-04	MA4AL1	10062	1.8597 -		SFAM01.1	PPB	11/17/2023 16:58	SMPL
CCV050	CCV050	25626	4.6941 -		SFAM01.1	PPB	11/17/2023 17:02	SMPL
CCB050	CCB050	-203	-0.0097 -		SFAM01.1	PPB	11/17/2023 17:05	SMPL