



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128373

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	11/17/2023	17:10
2	CAL2	1	Water	NA	NA	20.3	7.00	11/17/2023	17:11
3	CAL3	1	Water	NA	NA	20.6	10.02	11/17/2023	17:12
4	ICV	1	Water	NA	NA	20.5	7.01	11/17/2023	17:15
5	CCV1	1	Water	NA	NA	20.5	2.01	11/17/2023	17:20
6	O5481-06	1	Solid	20.02	20	23.2	7.50	11/17/2023	17:25
7	O5481-06DUP	1	Solid	20.03	20	23.3	7.51	11/17/2023	17:26
8	O5481-08	1	Solid	20.04	20	24.3	7.85	11/17/2023	17:30
9	O5481-09	1	Solid	20.02	20	24.1	7.38	11/17/2023	17:37
10	CCV2	1	Water	NA	NA	20.5	12.02	11/17/2023	17:40

128373

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S O5481 WorkList ID : 175678 Department : Wet-Chemistry Date : 11-17-2023 16:43:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5481-06	COMP-SOLID	Solid	pH	Cool 4 deg C	PSEG03		11/17/2023	9045D
O5481-08	60236-AQ	Solid	pH	Cool 4 deg C	PSEG03	L31	11/17/2023	9045D
O5481-09	60237-AQ	Solid	pH	Cool 4 deg C	PSEG03	L31	11/17/2023	9045D

Date/Time 11-17-23 16:50

Raw Sample Received by: JQ (CQC)

Raw Sample Relinquished by: JTC (CQC)

Date/Time 11-17-23 18:30
 Raw Sample Received by: JTC (CQC)
 Raw Sample Relinquished by: JQ (CQC)