


LB128377

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.0625e-004

C= 2.1625e-002

Rho= 0.9999433

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/v)
0.0	0.000	0.036	0.036	71	0.000	71	0				
0.2	0.200	0.212	0.012	924	0.0 %	924					6
2.5	2.500	2.485	-0.015	11945	0.0 %	11945					1
5.0	5.000	4.946	-0.054	23876	0.0 %	23876					1
7.5	7.500	7.463	-0.037	36078	0.0 %	36078					0
10.0	10.000	10.057	0.057	48658	0.0 %	48658					1

LB128377 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Nov 2023 10:53:37

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	71	-	0.0000	7470A	PPB	19 Nov 2023 11:29:15	S	Std
0.2 - 1	50-2	924	-	0.2000	7470A	PPB	19 Nov 2023 11:31:33	S	Std
2.5 - 1	52.5	11945	-	2.5000	7470A	PPB	19 Nov 2023 11:33:50	S	Std
5.0 - 1	55	23876	-	5.0000	7470A	PPB	19 Nov 2023 11:36:07	S	Std
7.5 - 1	57.5	36078	-	7.5000	7470A	PPB	19 Nov 2023 11:38:23	S	Std
10.0 - 1	510	48658	-	10.0000	7470A	PPB	19 Nov 2023 11:43:22	S	Std
ICV80 - 1	ICV80	18696	3.8777	-	7470A	PPB	19 Nov 2023 11:46:05	U	SMPL
ICB80 - 1	ICB80	-320	-0.0444	-	7470A	PPB	19 Nov 2023 11:48:21	U	SMPL
CCV51 - 1	CCV51	23624	4.8941	-	7470A	PPB	19 Nov 2023 11:50:38	U	SMPL
CCB51 - 1	CCB51	-315	-0.0433	-	7470A	PPB	19 Nov 2023 11:52:53	U	SMPL
CRA - 1	CRA	807	0.1881	-	7470A	PPB	19 Nov 2023 11:55:10	U	SMPL
HighStd - 1	HighStd	46685	9.6505	-	7470A	PPB	19 Nov 2023 11:57:25	U	SMPL
ChkStd - 1	ChkStd	33800	6.9929	-	7470A	PPB	19 Nov 2023 11:59:40	U	SMPL
PB157256BL - 1	PBW	-451	-0.0714	-	7470A	PPB	19 Nov 2023 12:01:59	U	SMPL
PB157256BS - 1	LCSW	20185	4.1848	-	7470A	PPB	19 Nov 2023 12:04:17	U	SMPL
O5445-03 - 1	SOIL-1-1115	-213	-0.0223	-	7470A	PPB	19 Nov 2023 12:06:33	U	SMPL
O5445-06 - 1	SOIL-2-1115	944	0.2163	-	7470A	PPB	19 Nov 2023 12:08:51	U	SMPL
O5445-09 - 1	SOIL-3-1115	-263	-0.0326	-	7470A	PPB	19 Nov 2023 12:11:08	U	SMPL
O5447-03 - 1	SP1	-20	0.0175	-	7470A	PPB	19 Nov 2023 12:13:26	U	SMPL
O5447-06 - 1	SP2	86	0.0394	-	7470A	PPB	19 Nov 2023 12:15:42	U	SMPL
CCV52 - 1	CCV52	23275	4.8221	-	7470A	PPB	19 Nov 2023 12:17:57	U	SMPL
CCB52 - 1	CCB52	-247	-0.0293	-	7470A	PPB	19 Nov 2023 12:20:13	U	SMPL
O5449-03 - 1	WC-SED-01	266	0.0765	-	7470A	PPB	19 Nov 2023 12:22:30	U	SMPL
O5449-06 - 1	WC-SED-02	719	0.1699	-	7470A	PPB	19 Nov 2023 12:24:45	U	SMPL
O5449-09 - 1	WC-SED-03	-11	0.0194	-	7470A	PPB	19 Nov 2023 12:27:01	U	SMPL
O5449-12 - 1	WC-SED-04	-47	0.0119	-	7470A	PPB	19 Nov 2023 12:29:18	U	SMPL
O5451-03 - 1	WC-SOIL-01	17	0.0251	-	7470A	PPB	19 Nov 2023 12:31:34	U	SMPL
O5462-01 - 1	S-21	1766	0.3859	-	7470A	PPB	19 Nov 2023 12:33:51	U	SMPL
O5462-02 - 1	S-22	791	0.1848	-	7470A	PPB	19 Nov 2023 12:36:07	U	SMPL
O5462-03 - 1	S-23	528	0.1305	-	7470A	PPB	19 Nov 2023 12:38:26	U	SMPL
O5462-04 - 1	S-24	367	0.0973	-	7470A	PPB	19 Nov 2023 12:40:44	U	SMPL
O5462-05 - 1	S-25	123	0.0470	-	7470A	PPB	19 Nov 2023 12:43:01	U	SMPL
CCV53 - 1	CCV53	24071	4.9863	-	7470A	PPB	19 Nov 2023 12:45:16	U	SMPL
CCB53 - 1	CCB53	-107	-0.0004	-	7470A	PPB	19 Nov 2023 12:47:38	U	SMPL
O5462-06 - 1	S-26	496	0.1239	-	7470A	PPB	19 Nov 2023 12:49:55	U	SMPL
O5462-07 - 1	S-27	122	0.0466	-	7470A	PPB	19 Nov 2023 12:52:11	U	SMPL
O5462-08 - 1	S-28	341	0.0920	-	7470A	PPB	19 Nov 2023 12:54:27	U	SMPL
O5466-03 - 1	RB21144	-49	0.0115	-	7470A	PPB	19 Nov 2023 12:56:43	U	SMPL
O5466-06 - 1	RB200018	29	0.0276	-	7470A	PPB	19 Nov 2023 12:58:59	U	SMPL
O5466-06DUP - 1	RB200018DUP	31	0.0280	-	7470A	PPB	19 Nov 2023 13:01:15	U	SMPL
O5466-06MS - 1	RB200018MS	18549	3.8474	-	7470A	PPB	19 Nov 2023 13:03:32	U	SMPL
O5466-06MSD - 1	RB200018MSD	16449	3.4143	-	7470A	PPB	19 Nov 2023 13:05:50	U	SMPL
PB157257BL - 1	PBW	-117	-0.0025	-	7470A	PPB	19 Nov 2023 13:08:07	U	SMPL
PB157257BS - 1	LCSW	21956	4.5501	-	7470A	PPB	19 Nov 2023 13:10:22	U	SMPL
CCV54 - 1	CCV54	23748	4.9197	-	7470A	PPB	19 Nov 2023 13:12:38	U	SMPL
CCB54 - 1	CCB54	-248	-0.0295	-	7470A	PPB	19 Nov 2023 13:14:54	U	SMPL
O5460-01 - 1	W-A-2-QUEENS-PLAZA	-47	0.0119	-	7470A	PPB	19 Nov 2023 13:17:11	U	SMPL
O5460-01DUP - 1	W-A-2-QUEENS-PLAZADUP	-31	0.0152	-	7470A	PPB	19 Nov 2023 13:19:28	U	SMPL
O5460-01MS - 1	W-A-2-QUEENS-PLAZAMS	20075	4.1621	-	7470A	PPB	19 Nov 2023 13:21:43	U	SMPL
O5460-02 - 1	W-A-8-50TH-AVE-11TH-ST	-314	-0.0431	-	7470A	PPB	19 Nov 2023 13:26:15	U	SMPL
O5460-03 - 1	W-A-3-JACKSON	-82	0.0047	-	7470A	PPB	19 Nov 2023 13:28:32	U	SMPL
O5460-04 - 1	W-A-3-HUNTER	-31	0.0152	-	7470A	PPB	19 Nov 2023 13:30:48	U	SMPL
O5460-05 - 1	W-A-4-HUNTER	25	0.0268	-	7470A	PPB	19 Nov 2023 13:33:06	U	SMPL
O5460-06 - 1	W-A-3-CORONA	6	0.0229	-	7470A	PPB	19 Nov 2023 13:35:22	U	SMPL
O5460-07 - 1	W-A-2-CORONA	7	0.0231	-	7470A	PPB	19 Nov 2023 13:37:38	U	SMPL
CCV55 - 1	CCV55	22480	4.6582	-	7470A	PPB	19 Nov 2023 13:39:54	U	SMPL
CCB55 - 1	CCB55	-232	-0.0262	-	7470A	PPB	19 Nov 2023 13:42:10	U	SMPL
PB157283BL - 1	PBW	-49	0.0115	-	7470A	PPB	19 Nov 2023 13:44:28	U	SMPL
PB157283BS - 1	LCSW	17939	3.7216	-	7470A	PPB	19 Nov 2023 13:46:43	U	SMPL
O5461-02 - 1	BATTERY-WASH-WATER	-266	-0.0332	-	7470A	PPB	19 Nov 2023 13:48:59	U	SMPL
O5461-02DUP - 1	BATTERY-WASH-WATERDUP	-64	0.0084	-	7470A	PPB	19 Nov 2023 13:51:15	U	SMPL
O5461-02MS - 1	BATTERY-WASH-WATERMS	17052	3.5386	-	7470A	PPB	19 Nov 2023 13:53:30	U	SMPL
O5461-02MSD - 1	BATTERY-WASH-WATERMSD	15983	3.3181	-	7470A	PPB	19 Nov 2023 13:55:46	U	SMPL
PB157200TB - 1	PB157200TB	-323	-0.0450	-	7470A	PPB	19 Nov 2023 13:58:05	U	SMPL
PB157229TB - 1	PB157229TB	-73	0.0066	-	7470A	PPB	19 Nov 2023 14:00:23	U	SMPL
PB157251TB - 1	PB157251TB	-33	0.0148	-	7470A	PPB	19 Nov 2023 14:02:41	U	SMPL
O5460-01MSD - 1		18902	3.9202	-	7470A	PPB	19 Nov 2023 14:04:57	U	SMPL
CCV56 - 1	CCV56	22399	4.6414	-	7470A	PPB	19 Nov 2023 14:07:13	U	SMPL
CCB56 - 1	CCB56	-339	-0.0483	-	7470A	PPB	19 Nov 2023 14:09:30	U	SMPL
O5466-06LX5 - 1		17	0.0251	-	7470A	PPB	19 Nov 2023 14:11:48	U	SMPL
O5466-06A - 1		18089	3.7525	-	7470A	PPB	19 Nov 2023 14:14:04	U	SMPL
O5460-01LX5 - 1		-160	-0.0114	-	7470A	PPB	19 Nov 2023 14:16:20	U	SMPL
O5460-01A - 1		17944	3.7226	-	7470A	PPB	19 Nov 2023 14:18:37	U	SMPL
O5461-02LX5 - 1		-134	-0.0060	-	7470A	PPB	19 Nov 2023 14:20:53	U	SMPL
O5461-02A - 1		19330	4.0085	-	7470A	PPB	19 Nov 2023 14:23:10	U	SMPL
CCV57 - 1	CCV57	22426	4.6470	-	7470A	PPB	19 Nov 2023 14:25:27	U	SMPL
CCB57 - 1	CCB57	-334	-0.0473	-	7470A	PPB	19 Nov 2023 14:27:45	U	SMPL