


LB128384

7471B

INSTRUMENT ID: 7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.0	0.000	0.008	0.008	305	0.000	305					-
0.2	0.200	0.172	-0.028	1171	0.0 %	1171					-9
2.5	2.500	2.449	-0.051	13158	0.0 %	13158					-2
5.0	5.000	5.066	0.066	26938	0.0 %	26938					-1
7.5	7.500	7.612	0.112	40349	0.0 %	40349					-1
10.0	10.000	9.893	-0.107	52360	0.0 %	52360					-1

LB128384 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Nov 2023 10:19:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	305	-	0.0000	7471B	PPB	20 Nov 2023 10:31:49	S	Std
0.2 - 1	50.2	1171	-	0.2000	7471B	PPB	20 Nov 2023 10:34:06	S	Std
2.5 - 1	52.5	13158	-	2.5000	7471B	PPB	20 Nov 2023 10:36:23	S	Std
5.0 - 1	55	26938	-	5.0000	7471B	PPB	20 Nov 2023 10:38:40	S	Std
7.5 - 1	57.5	40349	-	7.5000	7471B	PPB	20 Nov 2023 10:40:56	S	Std
10.0 - 1	510	52360	-	10.0000	7471B	PPB	20 Nov 2023 10:45:50	S	Std
ICV81 - 1	ICV81	21884	4.1058	-	7471B	PPB	20 Nov 2023 10:48:45	U	SMPL
ICB81 - 1	ICB81	-232	-0.0941	-	7471B	PPB	20 Nov 2023 10:51:01	U	SMPL
CCV58 - 1	CCV58	26609	5.0031	-	7471B	PPB	20 Nov 2023 10:53:18	U	SMPL
CCB58 - 1	CCB58	-224	-0.0926	-	7471B	PPB	20 Nov 2023 10:55:33	U	SMPL
CRA - 1	CRA	1275	0.1921	-	7471B	PPB	20 Nov 2023 11:05:06	U	SMPL
HighStd - 1	HighStd	51567	9.7426	-	7471B	PPB	20 Nov 2023 11:07:20	U	SMPL
ChkStd - 1	ChkStd	36719	6.9230	-	7471B	PPB	20 Nov 2023 11:09:35	U	SMPL
PB157284BL - 1	PB157284BL	280	0.0032	-	7471B	PPB	20 Nov 2023 11:14:37	U	SMPL
PB157284BS - 1	PB157284BS	22138	4.1540	-	7471B	PPB	20 Nov 2023 11:16:52	U	SMPL
O5434-01 - 1	COMP-1	3895	0.6896	-	7471B	PPB	20 Nov 2023 11:24:18	U	SMPL
O5434-01DUP - 1	COMP-1DUP	3551	0.6243	-	7471B	PPB	20 Nov 2023 11:29:10	U	SMPL
O5434-01MS - 1	COMP-1MS	24401	4.5838	-	7471B	PPB	20 Nov 2023 11:31:26	U	SMPL
O5434-01MSD - 1	COMP-1MSD	26565	4.9947	-	7471B	PPB	20 Nov 2023 11:33:43	U	SMPL
O5434-03 - 1	COMP-2	4174	0.7426	-	7471B	PPB	20 Nov 2023 11:36:01	U	SMPL
CCV59 - 1	CCV59	24119	4.5302	-	7471B	PPB	20 Nov 2023 11:38:18	U	SMPL
CCB59 - 1	CCB59	138	-0.0238	-	7471B	PPB	20 Nov 2023 11:40:33	U	SMPL
O5434-05 - 1	COMP-3	2585	0.4409	-	7471B	PPB	20 Nov 2023 11:42:50	U	SMPL
O5434-07 - 1	COMP-4	3686	0.6500	-	7471B	PPB	20 Nov 2023 11:45:05	U	SMPL
O5434-09 - 1	COMP-5	2641	0.4515	-	7471B	PPB	20 Nov 2023 11:47:21	U	SMPL
O5434-11 - 1	COMP-6	2253	0.3778	-	7471B	PPB	20 Nov 2023 11:49:36	U	SMPL
O5434-13 - 1	COMP-7	2660	0.4551	-	7471B	PPB	20 Nov 2023 11:51:52	U	SMPL
O5434-15 - 1	COMP-8	1110	0.1608	-	7471B	PPB	20 Nov 2023 11:54:09	U	SMPL
O5449-01 - 1	WC-SED-01	41822	7.8920	-	7471B	PPB	20 Nov 2023 11:56:25	U	SMPL
O5449-04 - 1	WC-SED-02	18395	3.4432	-	7471B	PPB	20 Nov 2023 11:58:42	U	SMPL
O5449-07 - 1	WC-SED-03	12328	2.2911	-	7471B	PPB	20 Nov 2023 12:01:02	U	SMPL
O5449-10 - 1	WC-SED-04	10698	1.9815	-	7471B	PPB	20 Nov 2023 12:03:19	U	SMPL
CCV60 - 1	CCV60	24212	4.5479	-	7471B	PPB	20 Nov 2023 12:05:36	U	SMPL
CCB60 - 1	CCB60	87	-0.0335	-	7471B	PPB	20 Nov 2023 12:07:53	U	SMPL
PB157285BL - 1	PB157285BL	296	0.0062	-	7471B	PPB	20 Nov 2023 12:11:47	U	SMPL
PB157285BS - 1	PB157285BS	23894	4.4875	-	7471B	PPB	20 Nov 2023 12:14:02	U	SMPL
O5406-01 - 1	OF-2-20231114	195645	37.1032	-	7471B	PPB	20 Nov 2023 12:16:17	U	SMPL
O5406-04 - 1	OF-3-20231114	1092	0.1574	-	7471B	PPB	20 Nov 2023 12:18:35	U	SMPL
O5406-04DUP - 1	OF-3-20231114DUP	1308	0.1984	-	7471B	PPB	20 Nov 2023 12:21:04	U	SMPL
O5406-05 - 1	OF-4-20231114	6517	1.1876	-	7471B	PPB	20 Nov 2023 12:23:20	U	SMPL
O5406-07 - 1	DUP-02-20231114	3225	0.5624	-	7471B	PPB	20 Nov 2023 12:25:37	U	SMPL
O5406-08 - 1	O5406-04MS	24839	4.6669	-	7471B	PPB	20 Nov 2023 12:27:54	U	SMPL
O5406-09 - 1	O5406-04MSD	25053	4.7076	-	7471B	PPB	20 Nov 2023 12:37:09	U	SMPL
O5451-01 - 1	WC-SOIL-01	2542	0.4327	-	7471B	PPB	20 Nov 2023 12:39:24	U	SMPL
CCV61 - 1	CCV61	28034	5.2737	-	7471B	PPB	20 Nov 2023 12:41:42	U	SMPL
CCB61 - 1	CCB61	155	-0.0206	-	7471B	PPB	20 Nov 2023 12:43:57	U	SMPL
O5463-01 - 1	NB-08-11162023	3345	0.5852	-	7471B	PPB	20 Nov 2023 12:46:14	U	SMPL
O5464-01 - 1	TR-05-11162023	2106	0.3499	-	7471B	PPB	20 Nov 2023 12:48:29	U	SMPL
O5466-01 - 1	RB21144	3611	0.6357	-	7471B	PPB	20 Nov 2023 12:50:45	U	SMPL
O5466-04 - 1	RB200018	3690	0.6507	-	7471B	PPB	20 Nov 2023 12:53:01	U	SMPL
O5468-01 - 1	TP-9	3723	0.6570	-	7471B	PPB	20 Nov 2023 12:55:16	U	SMPL
O5468-05 - 1	TP-10	1705	0.2738	-	7471B	PPB	20 Nov 2023 12:57:32	U	SMPL
O5468-09 - 1	TP-11	1067	0.1526	-	7471B	PPB	20 Nov 2023 12:59:48	U	SMPL
O5469-01 - 1	TP-12	1091	0.1572	-	7471B	PPB	20 Nov 2023 13:02:06	U	SMPL
CCV62 - 1	CCV62	28895	5.4372	-	7471B	PPB	20 Nov 2023 13:04:22	U	SMPL
CCB62 - 1	CCB62	295	0.0060	-	7471B	PPB	20 Nov 2023 13:06:38	U	SMPL
O5434-01LX5 - 1		813	0.1044	-	7471B	PPB	20 Nov 2023 13:08:54	U	SMPL
O5434-01A - 1		22947	4.3076	-	7471B	PPB	20 Nov 2023 13:11:09	U	SMPL
O5406-04LX5 - 1		234	-0.0056	-	7471B	PPB	20 Nov 2023 13:13:25	U	SMPL
O5406-04A - 1		25243	4.7437	-	7471B	PPB	20 Nov 2023 13:15:42	U	SMPL
O5406-01DLX5 - 1		42189	7.9617	-	7471B	PPB	20 Nov 2023 13:17:57	U	SMPL
CCV63 - 1	CCV63	28688	5.3979	-	7471B	PPB	20 Nov 2023 13:20:14	U	SMPL
CCB63 - 1	CCB63	177	-0.0164	-	7471B	PPB	20 Nov 2023 13:22:32	U	SMPL