



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB128387

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	11/17/2023	18:24
2	CAL2	1	Water	NA	NA	20.6	7.01	11/17/2023	18:25
3	CAL3	1	Water	NA	NA	20.1	10.02	11/17/2023	18:30
4	ICV	1	Water	NA	NA	20.3	6.99	11/17/2023	18:33
5	CCV1	1	Water	NA	NA	20.6	2.01	11/17/2023	18:35
6	O5461-02	1	Water	NA	NA	22.9	9.22	11/17/2023	18:36
7	O5481-01	1	Water	NA	NA	20.4	7.35	11/17/2023	18:40
8	O5481-04	1	Water	NA	NA	20.6	7.35	11/17/2023	18:41
9	O5481-05	1	Water	NA	NA	20.4	7.41	11/17/2023	18:44
10	O5484-01	1	Water	NA	NA	25.1	5.35	11/17/2023	18:45
11	O5484-02	1	Water	NA	NA	25.3	8.46	11/17/2023	18:50
12	O5486-01	1	Water	NA	NA	25.5	7.90	11/17/2023	19:00
13	O5486-01DUP	1	Water	NA	NA	25.6	7.91	11/17/2023	19:01
14	CCV2	1	Water	NA	NA	25.3	12.02	11/17/2023	19:05

WORKLIST(Hardcopy Internal Chain)

128387

WorkList Name : PH 05461

WorkList ID : 175675

Department : Wet-Chemistry

Date : 11-17-2023 13:41:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5461-02	BATTERY-WASH-WATER	Water	pH	Cool 4 deg C	ATCE02	L41	11/16/2023	9040C
O5481-01	COMP-AQ	Water	pH	Cool 4 deg C	PSEG03		11/17/2023	9040C
O5481-04	60236-AQ	Water	pH	Cool 4 deg C	PSEG03	L31	11/17/2023	9040C
O5481-05	60237-AQ	Water	pH	Cool 4 deg C	PSEG03	L31	11/17/2023	9040C
O5484-01	AUD-1292	Water	pH	Cool 4 deg C	PSEG03	L31	11/17/2023	9040C
O5484-02	AUD-1293	Water	pH	Cool 4 deg C	PSEG03	L31	11/17/2023	9040C
O5486-01	286118	Water	pH	Cool 4 deg C	PSEG03	L31	11/17/2023	9040C

Date/Time 11-17-23 18:20

Raw Sample Received by: 70160C

Raw Sample Relinquished by: 70160C

Date/Time 11-17-23

Raw Sample Received by: 70160C

Raw Sample Relinquished by: 70160C