



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128388

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	11/17/2023	18:24
2	CAL2	1	Water	NA	NA	20.6	7.01	11/17/2023	18:25
3	CAL3	1	Water	NA	NA	20.1	10.02	11/17/2023	18:30
4	ICV	1	Water	NA	NA	20.3	6.99	11/17/2023	18:33
5	CCV1	1	Water	NA	NA	20.5	2.01	11/17/2023	19:10
6	O5468-04	1	Solid	20.02	20	22.8	6.61	11/17/2023	19:15
7	O5468-08	1	Solid	20.03	20	22.7	6.47	11/17/2023	19:20
8	O5468-12	1	Solid	20.04	20	22.5	6.79	11/17/2023	19:22
9	O5469-04	1	Solid	20.02	20	22.2	6.59	11/17/2023	19:25
10	O5489-02	1	Solid	20.01	20	23.2	7.01	11/17/2023	19:30
11	O5489-05	1	Solid	20.03	20	23.5	7.38	11/17/2023	19:33
12	O5489-08	1	Solid	20.02	20	23.4	7.41	11/17/2023	19:35
13	O5492-04	1	Solid	20.02	20	24.1	7.56	11/17/2023	19:36
14	O5492-04DUP	1	Solid	20.03	20	24.2	7.57	11/17/2023	19:37
15	CCV2	1	Water	NA	NA	20.5	12.02	11/17/2023	19:39

WORKLIST(Hardcopy Internal Chain)

SP 128388

WorkList Name : CORRSOIVITY O54769 WorkList ID : 175674 Department : Wet-Chemistry Date : 11-17-2023 13:41:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5489-02	TRACKS-SP-1-E2	Solid	Corrosivity	Cool 4 deg C	TRAC01	L31	11/17/2023	9045D
O5489-05	TRACKS-SP-2-E2	Solid	Corrosivity	Cool 4 deg C	TRAC01	L31	11/17/2023	9045D
O5489-08	TRACKS-SP-3-E2	Solid	Corrosivity	Cool 4 deg C	TRAC01	L31	11/17/2023	9045D
O5468-04	TP-9	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	11/15/2023	9045D
O5468-08	TP-10	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	11/15/2023	9045D
O5468-12	TP-11	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	11/16/2023	9045D
O5469-04	TP-12	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	11/16/2023	9045D
O5492-04	TP-13	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	11/17/2023	9045D

Date/Time 11-17-23 18:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 11-17-23 19:50
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]