



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 11/21/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 11/20/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 11/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLIDS-OVEN

QC:LB128397

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5459-01	C0AH9	1	1.19	8.68	9.87	6.09	56.5	
O5459-02	C0AJ0	2	1.13	8.55	9.68	7.48	74.3	
O5459-03	VHBLK001	3	1.00	1.00	2.00	2.00	100.0	VHBLK
O5459-04	C0AA2	4	1.18	8.43	9.61	7.33	73.0	
O5459-05	C0AA3	5	1.15	8.51	9.66	7.23	71.4	
O5459-06	C0AA4	6	1.15	8.40	9.55	4.96	45.4	
O5459-07	C0AJ1	7	1.15	8.81	9.96	8.62	84.8	
O5459-08	C0AJ2	8	1.13	8.85	9.98	8.25	80.5	
O5459-09	C0AJ3	9	1.12	8.45	9.57	7.98	81.2	
O5459-10	C0AJ4	10	1.19	8.47	9.66	9.00	92.2	
O5459-11	C0AJ6	11	1.19	8.63	9.82	5.67	51.9	
O5459-12	C0AJ7	12	1.19	8.61	9.8	5.82	53.8	
O5459-13	C0AJ8	13	1.19	8.70	9.89	9.00	89.8	
O5459-14	C0AK0	14	1.19	8.52	9.71	8.44	85.1	
O5459-15	C0AK1	15	1.19	8.58	9.77	2.64	16.9	
O5459-16	C0AA0	16	1.15	8.80	9.95	7.61	73.4	
O5459-17	C0AA5	17	1.15	8.53	9.68	7.51	74.6	
O5459-18	C0AA6	18	1.14	8.38	9.52	6.21	60.5	
O5459-19	C0AA7	19	1.16	8.80	9.96	7.26	69.3	
O5459-20	C0AA8	20	1.12	8.65	9.77	7.44	73.1	
O5459-21	C0AA9	21	1.19	8.72	9.91	7.63	73.9	
O5459-22	C0AA9MS	22	1.19	8.72	9.91	7.63	73.9	
O5459-23	C0AA9MSD	23	1.19	8.72	9.91	7.63	73.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$