



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128398

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.4	11/20/2023	12:00
2	O5481-01		1	12	100.00	>212.0	>212.0	11/20/2023	12:30
3	O5482-01		1	14	100.00	>212.0	>212.0	11/20/2023	13:00
4	O5482-02		1	12	100.00	>212.0	>212.0	11/20/2023	13:30
5	O5486-01		1	14	100.00	>212.0	>212.0	11/20/2023	14:00
6	O5486-01DUP		1	14	100.00	>212.0	>212.0	11/20/2023	14:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6128398

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp11-20 WorkList ID : 175705 Department : Wet-Chemistry Date : 11-20-2023 08:35:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5481-01	COMP-AQ	Water	Flash Point	Cool 4 deg C	PSEG03	L31	11/17/2023	1010B
O5482-01	1142	Water	Flash Point	Cool 4 deg C	PSEG03	L41	11/17/2023	1010B
O5482-02	1141	Water	Flash Point	Cool 4 deg C	PSEG03	L41	11/17/2023	1010B
O5486-01	286118	Water	Flash Point	Cool 4 deg C	PSEG03	L31	11/17/2023	1010B

Date/Time 11/20/2023 11:53
Raw Sample Received by: RM 11/20/2023
Raw Sample Relinquished by: 11/20/2023

Date/Time 11/20/2023 15:10
Raw Sample Received by: 3d Coc 11/20/2023
Raw Sample Relinquished by: Rpt 11/20/2023