

LB128406


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

4988E



A= 0.0000e+000

B= 1.9929e-004 slope

C= 3.2959e-002 y-intercept

Rho= 0.9998074

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.00	0.000	0.034	0.034	7	0.000	7					
0.05	0.050					419	0				
0.20	0.200	0.255	0.055	1116	0.0 %	1116					
2.50	2.500	2.394	-0.106	11846	0.0 %	11846					
5.00	5.000	4.936	-0.064	24600	0.0 %	24600					
7.50	7.500	7.606	0.106	37998	0.0 %	37998					
10.0	10.000	9.975	-0.025	49888	0.0 %	49888					

LB128406

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	7	-		0 SFAM01.1	PPB	11/21/2023 10:44	Std
	0.2 S01	1116	-		0.2 SFAM01.1	PPB	11/21/2023 10:50	Std
	2.5 S02	11846	-		2.5 SFAM01.1	PPB	11/21/2023 10:52	Std
	5 S03	24600	-		5 SFAM01.1	PPB	11/21/2023 10:54	Std
	7.5 S04	37998	-		7.5 SFAM01.1	PPB	11/21/2023 10:57	Std
	10 S05	49888	-		10 SFAM01.1	PPB	11/21/2023 10:59	Std
ICV083	ICV083	20837	4.1856 -		SFAM01.1	PPB	11/21/2023 11:02	SMPL
ICB083	ICB083	-199	-0.0067 -		SFAM01.1	PPB	11/21/2023 11:04	SMPL
CCV066	CCV066	25431	5.1012 -		SFAM01.1	PPB	11/21/2023 11:06	SMPL
CCB066	CCB066	-216	-0.0101 -		SFAM01.1	PPB	11/21/2023 11:09	SMPL
PB157317BL	PBW317	8	0.0346 -		SFAM01.1	PPB	11/21/2023 11:11	SMPL
O5240-01	MA0BB6	33	0.0395 -		SFAM01.1	PPB	11/21/2023 11:13	SMPL
O5240-02	MA0BB6D	51	0.0431 -		SFAM01.1	PPB	11/21/2023 11:15	SMPL
O5240-03	MA0BB6S	5647	1.1584 -		SFAM01.1	PPB	11/21/2023 11:18	SMPL
O5240-04	MA0BB7	0	0.033 -		SFAM01.1	PPB	11/21/2023 11:20	SMPL
O5240-05	MA0BC1	7257	1.4792 -		SFAM01.1	PPB	11/21/2023 11:22	SMPL
O5240-06	MA0BC2	41956	8.3945 -		SFAM01.1	PPB	11/21/2023 11:29	SMPL
CCV067	CCV067	23981	4.8122 -		SFAM01.1	PPB	11/21/2023 11:38	SMPL
CCB067	CCB067	29	0.0387 -		SFAM01.1	PPB	11/21/2023 11:41	SMPL