



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128411

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/21/2023	13:20
2	CAL2	1	Water	NA	NA	20.6	7.00	11/21/2023	13:21
3	CAL3	1	Water	NA	NA	20.5	10.02	11/21/2023	13:22
4	ICV	1	Water	NA	NA	20.3	7.02	11/21/2023	13:25
5	CCV1	1	Water	NA	NA	20.5	2.01	11/21/2023	13:30
6	O5510-01	1	Water	NA	NA	20.6	4.58	11/21/2023	13:35
7	O5510-01DUP	1	Water	NA	NA	20.5	4.59	11/21/2023	13:36
8	O5510-02	1	Water	NA	NA	20.7	4.87	11/21/2023	13:40
9	O5510-03	1	Water	NA	NA	20.6	4.83	11/21/2023	13:45
10	CCV2	1	Water	NA	NA	20.3	12.02	11/21/2023	13:50

W2128411

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o5510 WorkList ID : 175756 Department : Wet-Chemistry Date : 11-21-2023 12:55:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5510-01	A CELL-A-POOL-WATER	Water	pH	Cool 4 deg C	STER13	L31	11/20/2023	9040C
O5510-02	A CELL-B-POOL-WATER	Water	pH	Cool 4 deg C	STER13	L31	11/20/2023	9040C
O5510-03	A CELL-C-POOL-WATER	Water	pH	Cool 4 deg C	STER13	L31	11/20/2023	9040C

Date/Time 11/21/23 13:10
 Raw Sample Received by: JF
 Raw Sample Relinquished by: JF

Date/Time 11/21/23 14:30
 Raw Sample Received by: JF
 Raw Sample Relinquished by: JF