

LB128412


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50877



A= 0.0000e+000

B= 1.9468e-004

C= -1.2094e-002

Rho= 0.9995889

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.053	0.053	332	0.000	332					
0.05	0.050					419	0				11
0.20	0.200	0.210	0.010	1141	0.0 %	1141					5
2.50	2.500	2.383	-0.117	12304	0.0 %	12304					-5
5.00	5.000	4.966	-0.034	25571	0.0 %	25571					-1
7.50	7.500	7.695	0.195	39588	0.0 %	39588					3
10.0	10.000	9.893	-0.107	50877	0.0 %	50877					-1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	332	-		0 SFAM01.1	PPB	11/21/2023 14:31	Std
	0.2 S01	1141	-		0.2 SFAM01.1	PPB	11/21/2023 14:33	Std
	2.5 S02	12304	-		2.5 SFAM01.1	PPB	11/21/2023 14:36	Std
	5 S03	25571	-		5 SFAM01.1	PPB	11/21/2023 14:38	Std
	7.5 S04	39588	-		7.5 SFAM01.1	PPB	11/21/2023 14:40	Std
	10 S05	50877	-		10 SFAM01.1	PPB	11/21/2023 14:42	Std
ICV84	ICV84	20900	4.0568 -		SFAM01.1	PPB	11/21/2023 14:45	SMPL
ICB84	ICB84	28	-0.0066 -		SFAM01.1	PPB	11/21/2023 14:48	SMPL
CCV68	CCV68	25636	4.9788 -		SFAM01.1	PPB	11/21/2023 14:52	SMPL
CCB68	CCB68	49	-0.0026 -		SFAM01.1	PPB	11/21/2023 14:55	SMPL
PB157346BL	PBW346	82	0.0039 -		SFAM01.1	PPB	11/21/2023 14:57	SMPL
PB156939TB	LEB346	132	0.0136 -		SFAM01.1	PPB	11/21/2023 14:59	SMPL
PB157311TB	LEB346	136	0.0144 -		SFAM01.1	PPB	11/21/2023 15:01	SMPL
O5204-01	MCOAN8	156	0.0183 -		SFAM01.1	PPB	11/21/2023 15:04	SMPL
O5204-02	MCOAN8D	170	0.021 -		SFAM01.1	PPB	11/21/2023 15:06	SMPL
O5204-03	MCOAN8S	5163	0.9931 -		SFAM01.1	PPB	11/21/2023 15:08	SMPL
O5479-01	MJLAD9	4782	0.9189 -		SFAM01.1	PPB	11/21/2023 15:11	SMPL
O5479-02	MJLDA8	398	0.0654 -		SFAM01.1	PPB	11/21/2023 15:20	SMPL
O5479-03	MJLDA8D	326	0.0514 -		SFAM01.1	PPB	11/21/2023 15:23	SMPL
O5479-04	MJLDA8S	5006	0.9625 -		SFAM01.1	PPB	11/21/2023 15:25	SMPL
CCV069	CCV069	22674	4.4022 -		SFAM01.1	PPB	11/21/2023 15:27	SMPL
CCB069	CCB069	188	0.0245 -		SFAM01.1	PPB	11/21/2023 15:29	SMPL