

L3128421


7470A

INSTRUMENT ID: CY1

Linear

μ Abs.:

38128



A= 0.0000e+000

B= 2.5737e-004

C= -1.5016e-001

Rho= 0.9983173

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	-0.088	-0.088	241	0.000	241	0				
0.2	0.200	0.217	0.017	1425	0.0 %	1425					
2.5	2.500	2.387	-0.113	9860	0.0 %	9860					
5.0	5.000	5.214	0.214	20841	0.0 %	20841					
7.5	7.500	7.808	0.308	30920	0.0 %	30920					
10.0	10.000	9.663	-0.337	38128	0.0 %	38128					

1.8
1.5
1.4
1.4
1.3

LB128421 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Nov 2023 08:45:52

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	241	-	0.0000	7470A	PPB	22 Nov 2023 08:59:28	S	Std
0.2 - 1	50.2	1425	-	0.2000	7470A	PPB	22 Nov 2023 09:01:45	S	Std
2.5 - 1	52.5	9860	-	2.5000	7470A	PPB	22 Nov 2023 09:04:02	S	Std
5.0 - 1	55	20841	-	5.0000	7470A	PPB	22 Nov 2023 09:06:18	S	Std
7.5 - 1	57.5	30920	-	7.5000	7470A	PPB	22 Nov 2023 09:08:34	S	Std
10.0 - 1	510	38128	-	10.0000	7470A	PPB	22 Nov 2023 09:15:55	S	Std
ICV85 - 1	ICV85	16611	4.1250	-	7470A	PPB	22 Nov 2023 09:18:42	U	SMPL
ICB85 - 1	ICB85	518	-0.0168	-	7470A	PPB	22 Nov 2023 09:23:33	U	SMPL
CCV70 - 1	CCV70	20233	5.0572	-	7470A	PPB	22 Nov 2023 09:25:48	U	SMPL
CCB70 - 1	CCB70	464	-0.0307	-	7470A	PPB	22 Nov 2023 09:28:03	U	SMPL
CRA - 1	CRA	1233	0.1672	-	7470A	PPB	22 Nov 2023 09:30:25	U	SMPL
HighStd - 1	HighStd	35914	9.0929	-	7470A	PPB	22 Nov 2023 09:35:20	U	SMPL
ChkStd - 1	ChkStd	26785	6.7434	-	7470A	PPB	22 Nov 2023 09:37:36	U	SMPL
PB157312BL - 1	PBW	450	-0.0343	-	7470A	PPB	22 Nov 2023 09:42:30	U	SMPL
PB157312BS - 1	LCSW	15960	3.9574	-	7470A	PPB	22 Nov 2023 09:44:46	U	SMPL
O5468-04 - 1	TP-9	1359	0.1996	-	7470A	PPB	22 Nov 2023 09:47:03	U	SMPL
O5468-08 - 1	TP-10	571	-0.0032	-	7470A	PPB	22 Nov 2023 09:49:20	U	SMPL
O5468-12 - 1	TP-11	176	-0.1049	-	7470A	PPB	22 Nov 2023 09:51:39	U	SMPL
O5469-04 - 1	TP-12	769	0.0478	-	7470A	PPB	22 Nov 2023 09:53:57	U	SMPL
O5485-02 - 1	AU-14-STOCKPILE-BIN-5	770	0.0480	-	7470A	PPB	22 Nov 2023 09:56:12	U	SMPL
CCV71 - 1	CCV71	20323	5.0803	-	7470A	PPB	22 Nov 2023 09:58:28	U	SMPL
CCB71 - 1	CCB71	436	-0.0379	-	7470A	PPB	22 Nov 2023 10:00:44	U	SMPL
O5489-02 - 1	TRACKS-SP-1-E2	692	0.0279	-	7470A	PPB	22 Nov 2023 10:03:00	U	SMPL
O5489-05 - 1	TRACKS-SP-2-E2	645	0.0158	-	7470A	PPB	22 Nov 2023 10:05:15	U	SMPL
O5489-08 - 1	TRACKS-SP-3-E2	747	0.0421	-	7470A	PPB	22 Nov 2023 10:07:31	U	SMPL
O5492-04 - 1	TP-13	784	0.0516	-	7470A	PPB	22 Nov 2023 10:09:47	U	SMPL
O5492-04DUP - 1	TP-13DUP	782	0.0511	-	7470A	PPB	22 Nov 2023 10:12:03	U	SMPL
O5492-04MS - 1	TP-13MS	15991	3.9654	-	7470A	PPB	22 Nov 2023 10:14:20	U	SMPL
O5492-04MSD - 1	TP-13MSD	17210	4.2791	-	7470A	PPB	22 Nov 2023 10:31:04	U	SMPL
PB157371BL - 1	PBW	558	-0.0065	-	7470A	PPB	22 Nov 2023 10:33:21	U	SMPL
PB157371BS - 1	LCSW	17858	4.4459	-	7470A	PPB	22 Nov 2023 10:39:13	U	SMPL
O5425-05 - 1	EB20231114	909	0.0838	-	7470A	PPB	22 Nov 2023 10:41:28	U	SMPL
CCV72 - 1	CCV72	21032	5.2628	-	7470A	PPB	22 Nov 2023 10:43:44	U	SMPL
CCB72 - 1	CCB72	523	-0.0156	-	7470A	PPB	22 Nov 2023 10:46:00	U	SMPL
O5425-06 - 1	FB20231114	929	0.0889	-	7470A	PPB	22 Nov 2023 10:48:17	U	SMPL
O5481-01 - 1	COMP-AQ	91697	23.4497	-	7470A	PPB	22 Nov 2023 10:50:33	U	SMPL
O5482-02 - 1	1141	4853	1.0988	-	7470A	PPB	22 Nov 2023 10:52:48	U	SMPL
O5484-01 - 1	AUD-1292	2540	0.5036	-	7470A	PPB	22 Nov 2023 10:55:07	U	SMPL
O5484-01DUP - 1	AUD-1292DUP	2564	0.5097	-	7470A	PPB	22 Nov 2023 10:57:24	U	SMPL
O5484-01MS - 1	AUD-1292MS	19320	4.8222	-	7470A	PPB	22 Nov 2023 10:59:40	U	SMPL
O5484-01MSD - 1	AUD-1292MSD	16459	4.0859	-	7470A	PPB	22 Nov 2023 11:01:57	U	SMPL
O5486-01 - 1	286118	1802	0.3136	-	7470A	PPB	22 Nov 2023 11:04:15	U	SMPL
CCV73 - 1	CCV73	21517	5.3876	-	7470A	PPB	22 Nov 2023 11:09:26	U	SMPL
CCB73 - 1	CCB73	602	0.0048	-	7470A	PPB	22 Nov 2023 11:11:41	U	SMPL
PB157249TB - 1		585	0.0004	-	7470A	PPB	22 Nov 2023 11:13:57	U	SMPL
O5492-04LX5 - 1		250	-0.0858	-	7470A	PPB	22 Nov 2023 11:16:12	U	SMPL
O5492-04A - 1		16732	4.1561	-	7470A	PPB	22 Nov 2023 11:18:28	U	SMPL
O5484-01LX5 - 1		1093	0.1311	-	7470A	PPB	22 Nov 2023 11:32:02	U	SMPL
O5484-01A - 1		6323	1.4772	-	7470A	PPB	22 Nov 2023 11:34:17	U	SMPL
O5481-01DLX5 - 1		21459	5.3727	-	7470A	PPB	22 Nov 2023 11:41:42	U	SMPL
CCV74 - 1	CCV74	19682	4.9154	-	7470A	PPB	22 Nov 2023 11:43:58	U	SMPL
CCB74 - 1	CCB74	291	-0.0753	-	7470A	PPB	22 Nov 2023 11:46:14	U	SMPL