


LB128429

7471B

INSTRUMENT ID: CV3

Linear



A= 0.0000e+000

B= 1.4444e-004

C= 4.6715e-002

Rho= 0.9997675

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.064	0.064	117	0.000	117					0
0.2	0.200	0.220	0.020	1202	0.0 %	1202					10
2.5	2.500	2.338	-0.162	15866	0.0 %	15866					-6
5.0	5.000	5.086	0.086	34887	0.0 %	34887					2
7.5	7.500	7.494	-0.006	51558	0.0 %	51558					0
10.0	10.000	9.998	-0.002	68893	0.0 %	68893					0

LB128429 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Nov 2023 13:13:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	117	-	0.0000	7471B	PPB	22 Nov 2023 13:26:44	S	Std
0.2 - 1	50.2	1202	-	0.2000	7471B	PPB	22 Nov 2023 13:29:01	S	Std
2.5 - 1	52.5	15866	-	2.5000	7471B	PPB	22 Nov 2023 13:31:18	S	Std
5.0 - 1	55	34887	-	5.0000	7471B	PPB	22 Nov 2023 13:33:35	S	Std
7.5 - 1	57.5	51558	-	7.5000	7471B	PPB	22 Nov 2023 13:35:51	S	Std
10.0 - 1	510	68893	-	10.0000	7471B	PPB	22 Nov 2023 13:38:09	S	Std
ICV86 - 1	ICV86	25049	3.6649	-	7471B	PPB	22 Nov 2023 13:41:02	U	SMPL
ICB86 - 1	ICB86	-210	0.0164	-	7471B	PPB	22 Nov 2023 13:43:17	U	SMPL
CCV75 - 1	CCV75	34965	5.0972	-	7471B	PPB	22 Nov 2023 13:45:35	U	SMPL
CCB75 - 1	CCB75	-222	0.0146	-	7471B	PPB	22 Nov 2023 13:47:49	U	SMPL
CRA - 1	CRA	1161	0.2144	-	7471B	PPB	22 Nov 2023 13:50:06	U	SMPL
HighStd - 1	HighStd	69242	10.0482	-	7471B	PPB	22 Nov 2023 13:52:21	U	SMPL
ChkStd - 1	ChkStd	51602	7.5003	-	7471B	PPB	22 Nov 2023 13:57:27	U	SMPL
PB157387BL - 1	PBS	-356	-0.0047	-	7471B	PPB	22 Nov 2023 13:59:42	U	SMPL
PB157387BS - 1	LCSS	25325	3.7047	-	7471B	PPB	22 Nov 2023 14:04:35	U	SMPL
O5485-01 - 1	AU-14-STOCKPILE-BIN-5	6359	0.9652	-	7471B	PPB	22 Nov 2023 14:06:51	U	SMPL
O5489-01 - 1	TRACKS-SP-1	24340	3.5625	-	7471B	PPB	22 Nov 2023 14:09:09	U	SMPL
O5489-04 - 1	TRACKS-SP-2	33628	4.9040	-	7471B	PPB	22 Nov 2023 14:11:27	U	SMPL
O5489-07 - 1	TRACKS-SP-3	34594	5.0436	-	7471B	PPB	22 Nov 2023 14:13:45	U	SMPL
O5492-01 - 1	TP-13	13122	1.9421	-	7471B	PPB	22 Nov 2023 14:16:02	U	SMPL
CCV76 - 1	CCV76	35077	5.1133	-	7471B	PPB	22 Nov 2023 14:18:20	U	SMPL
CCB76 - 1	CCB76	-122	0.0291	-	7471B	PPB	22 Nov 2023 14:20:35	U	SMPL
O5503-02 - 1	358	32512	4.7428	-	7471B	PPB	22 Nov 2023 14:22:51	U	SMPL
O5503-04 - 1	72-11998	7999	1.2021	-	7471B	PPB	22 Nov 2023 14:25:07	U	SMPL
O5504-02 - 1	281	8832	1.3224	-	7471B	PPB	22 Nov 2023 14:27:23	U	SMPL
O5505-01 - 1	TP-A	709	0.1491	-	7471B	PPB	22 Nov 2023 14:29:39	U	SMPL
O5505-04 - 1	TP-D	248	0.0825	-	7471B	PPB	22 Nov 2023 14:31:55	U	SMPL
O5507-01 - 1	SU-04-11202023	11970	1.7757	-	7471B	PPB	22 Nov 2023 14:39:09	U	SMPL
O5507-01DUP - 1	SU-04-11202023DUP	13447	1.9890	-	7471B	PPB	22 Nov 2023 14:41:25	U	SMPL
O5507-01MS - 1	SU-04-11202023MS	32458	4.7350	-	7471B	PPB	22 Nov 2023 14:43:42	U	SMPL
O5507-01MSD - 1	SU-04-11202023MSD	36426	5.3082	-	7471B	PPB	22 Nov 2023 14:46:00	U	SMPL
CCV77 - 1	CCV77	0	0.0467	-	7471B	PPB	22 Nov 2023 14:48:16	U	SMPL
O5507-01MS - 1	SU-04-11202023MS	34688	5.0571	-	7471B	PPB	22 Nov 2023 14:48:34	U	SMPL
O5507-01MSD - 1	SU-04-11202023MSD	34534	5.0349	-	7471B	PPB	22 Nov 2023 14:50:51	U	SMPL
CCV77 - 1	CCV77	36104	5.2617	-	7471B	PPB	22 Nov 2023 14:53:08	U	SMPL
CCB77 - 1	CCB77	-50	0.0395	-	7471B	PPB	22 Nov 2023 14:55:24	U	SMPL
O5507-01LX5 - 1		2628	0.4263	-	7471B	PPB	22 Nov 2023 14:57:40	U	SMPL
O5507-01A - 1		39380	5.7349	-	7471B	PPB	22 Nov 2023 14:59:56	U	SMPL
CCV78 - 1	CCV78	35207	5.1321	-	7471B	PPB	22 Nov 2023 15:02:12	U	SMPL
CCB78 - 1	CCB78	-69	0.0367	-	7471B	PPB	22 Nov 2023 15:04:28	U	SMPL