

LB128451


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53568



A= 0.0000e+000

B= 1.8775e-004 slope

C= -3.2448e-002 intercept

Rho= 0.9999605

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.00	0.000	-0.028	-0.028	24	0.000	24					
0.05	0.050					419	0				
0.20	0.200	0.205	0.005	1263	0.0 %	1263					
2.50	2.500	2.547	0.047	13737	0.0 %	13737					
5.00	5.000	5.004	0.004	26827	0.0 %	26827					
7.50	7.500	7.447	-0.053	39836	0.0 %	39836					
10.0	10.000	10.025	0.025	53568	0.0 %	53568					

0.013
11
3
2
0
1
0

INSTRUMENT IS : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	24	-		0 SFAM01.1	PPB	11/24/2023 13:15	Std
	0.2 S01	1263	-		0.2 SFAM01.1	PPB	11/24/2023 13:18	Std
	2.5 S02	13737	-		2.5 SFAM01.1	PPB	11/24/2023 13:21	Std
	5 S03	26827	-		5 SFAM01.1	PPB	11/24/2023 13:24	Std
	7.5 S04	39836	-		7.5 SFAM01.1	PPB	11/24/2023 13:26	Std
	10 S05	53568	-		10 SFAM01.1	PPB	11/24/2023 13:28	Std
ICV087	ICV087	21571	4.0176 -		SFAM01.1	PPB	11/24/2023 13:31	SMPL
ICB087	ICB087	-175	-0.0653 -		SFAM01.1	PPB	11/24/2023 13:34	SMPL
CCV078	CCV078	26691	4.9789 -		SFAM01.1	PPB	11/24/2023 13:36	SMPL
CCB078	CCB078	-269	-0.083 -		SFAM01.1	PPB	11/24/2023 13:38	SMPL
PB157402BL	PBS402	-32	-0.0385 -		SFAM01.1	PPB	11/24/2023 13:40	SMPL
O5175-01	MCOAM5	508	0.0629 -		SFAM01.1	PPB	11/24/2023 13:43	SMPL
O5175-02	MCOAM6	-45	-0.0409 -		SFAM01.1	PPB	11/24/2023 13:45	SMPL
O5175-03	MCOAN1	2494	0.4358 -		SFAM01.1	PPB	11/24/2023 13:47	SMPL
O5175-04	MCOAN2	-73	-0.0462 -		SFAM01.1	PPB	11/24/2023 13:49	SMPL
O5175-05	MCOAN6	1660	0.2792 -		SFAM01.1	PPB	11/24/2023 13:57	SMPL
O5175-06	MCOAN6D	1609	0.2696 -		SFAM01.1	PPB	11/24/2023 13:59	SMPL
O5175-07	MCOAN6S	13579	2.5171 -		SFAM01.1	PPB	11/24/2023 14:02	SMPL
O5175-08	MCOAN7	5	-0.0315 -		SFAM01.1	PPB	11/24/2023 14:04	SMPL
O5175-09	MCOAP2	298	0.0235 -		SFAM01.1	PPB	11/24/2023 14:06	SMPL
O5175-10	MCOAP9	499	0.0612 -		SFAM01.1	PPB	11/24/2023 14:08	SMPL
O5175-11	MCOAQ0	54	-0.0223 -		SFAM01.1	PPB	11/24/2023 14:11	SMPL
O5175-12	MCOAQ1	60	-0.0212 -		SFAM01.1	PPB	11/24/2023 14:13	SMPL
CCV079	CCV079	26543	4.9511 -		SFAM01.1	PPB	11/24/2023 14:34	SMPL
CCB079	CCB079	-141	-0.0589 -		SFAM01.1	PPB	11/24/2023 14:36	SMPL