

LB128458


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

55062



A= 0.0000e+000

B= 1.7695e-004

C= -2.3459e-002

Rho= 0.9987429

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.0	0.000	-0.042	-0.042	-102	0.000	-102					
0.2	0.200	0.217	0.017	1359	0.0 %	1359					
2.5	2.500	2.367	-0.133	13509	0.0 %	13509					
5.0	5.000	5.143	0.143	29196	0.0 %	29196					
7.5	7.500	7.795	0.295	44181	0.0 %	44181					
10.0	10.000	9.720	-0.280	55062	0.0 %	55062					

101
181
13

LB128458 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Nov 2023 14:43:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-102	-	0.0000	7471B	PPB	27 Nov 2023 14:45:57	S	Std
0.2 - 1	50-2	1359	-	0.2000	7471B	PPB	27 Nov 2023 14:48:14	S	Std
2.5 - 1	52.5	13509	-	2.5000	7471B	PPB	27 Nov 2023 14:50:31	S	Std
5.0 - 1	55	29196	-	5.0000	7471B	PPB	27 Nov 2023 14:52:48	S	Std
7.5 - 1	57.5	44181	-	7.5000	7471B	PPB	27 Nov 2023 14:55:07	S	Std
10.0 - 1	510	55062	-	10.0000	7471B	PPB	27 Nov 2023 15:00:13	S	Std
ICV89 - 1	ICV89	22026	3.8741	-	7471B	PPB	27 Nov 2023 15:03:04	U	SMPL
ICB89 - 1	ICB89	-85	-0.0385	-	7471B	PPB	27 Nov 2023 15:05:22	U	SMPL
CCV84 - 1	CCV84	29844	5.2576	-	7471B	PPB	27 Nov 2023 15:07:43	U	SMPL
CCB84 - 1	CCB84	-86	-0.0387	-	7471B	PPB	27 Nov 2023 15:09:58	U	SMPL
CRA - 1	CRA	1341	0.2138	-	7471B	PPB	27 Nov 2023 15:12:22	U	SMPL
HighStd - 1	HighStd	55511	9.7995	-	7471B	PPB	27 Nov 2023 15:14:36	U	SMPL
ChkStd - 1	ChkStd	41561	7.3309	-	7471B	PPB	27 Nov 2023 15:16:51	U	SMPL
PB157422BL - 1	PBS	-345	-0.0845	-	7471B	PPB	27 Nov 2023 15:19:22	U	SMPL
PB157422BS - 1	LCSS	21926	3.8564	-	7471B	PPB	27 Nov 2023 15:21:48	U	SMPL
O5516-01 - 1	72-11930	6030	1.0436	-	7471B	PPB	27 Nov 2023 15:24:04	U	SMPL
O5516-03 - 1	ETGI-343	2441	0.4085	-	7471B	PPB	27 Nov 2023 15:26:24	U	SMPL
O5517-01 - 1	TP-A	954	0.1454	-	7471B	PPB	27 Nov 2023 15:28:43	U	SMPL
O5517-04 - 1	TP-D	888	0.1337	-	7471B	PPB	27 Nov 2023 15:31:01	U	SMPL
O5518-01 - 1	31923	1607	0.2609	-	7471B	PPB	27 Nov 2023 15:33:17	U	SMPL
CCV85 - 1	CCV85	26703	4.7018	-	7471B	PPB	27 Nov 2023 15:35:32	U	SMPL
CCB85 - 1	CCB85	465	0.0588	-	7471B	PPB	27 Nov 2023 15:42:55	U	SMPL
O5518-02 - 1	1766	104	-0.0051	-	7471B	PPB	27 Nov 2023 15:45:10	U	SMPL
O5518-03 - 1	31921	2536	0.4253	-	7471B	PPB	27 Nov 2023 15:47:25	U	SMPL
O5518-04 - 1	1109	85862	15.1702	-	7471B	PPB	27 Nov 2023 15:49:41	U	SMPL
O5518-05 - 1	1110	147097	26.0060	-	7471B	PPB	27 Nov 2023 15:51:59	U	SMPL
O5519-01 - 1	OK-02-112123	-227	-0.0636	-	7471B	PPB	27 Nov 2023 15:54:34	U	SMPL
O5521-01 - 1	SP-1-COMP	7099	1.2327	-	7471B	PPB	27 Nov 2023 15:57:14	U	SMPL
O5521-01DUP - 1	SP-1-COMPDUP	7595	1.3205	-	7471B	PPB	27 Nov 2023 15:59:30	U	SMPL
O5521-01MS - 1	SP-1-COMPMS	28981	5.1049	-	7471B	PPB	27 Nov 2023 16:01:50	U	SMPL
O5521-01MSD - 1	SP-1-COMPMSD	25588	4.5045	-	7471B	PPB	27 Nov 2023 16:04:10	U	SMPL
O5521-04 - 1	SP-2-COMP	9056	1.5790	-	7471B	PPB	27 Nov 2023 16:06:33	U	SMPL
CCV86 - 1	CCV86	26908	4.7380	-	7471B	PPB	27 Nov 2023 16:08:57	U	SMPL
CCB86 - 1	CCB86	-5	-0.0243	-	7471B	PPB	27 Nov 2023 16:11:15	U	SMPL
O5521-07 - 1	SP-3-COMP	8933	1.5573	-	7471B	PPB	27 Nov 2023 16:13:38	U	SMPL
O5521-10 - 1	SP-4-COMP	8709	1.5176	-	7471B	PPB	27 Nov 2023 16:15:53	U	SMPL
O5539-01 - 1	EO-03-112223	2075	0.3437	-	7471B	PPB	27 Nov 2023 16:18:13	U	SMPL
O5540-01 - 1	TR-06-11222023	1609	0.2613	-	7471B	PPB	27 Nov 2023 16:20:31	U	SMPL
O5521-01LX5 - 1		1296	0.2059	-	7471B	PPB	27 Nov 2023 16:25:27	U	SMPL
O5521-01A - 1		28996	5.1075	-	7471B	PPB	27 Nov 2023 16:27:43	U	SMPL
O5518-04DLX5 - 1		17976	3.1575	-	7471B	PPB	27 Nov 2023 16:30:00	U	SMPL
O5518-05DLX5 - 1		34706	6.1179	-	7471B	PPB	27 Nov 2023 16:32:24	U	SMPL
CCV87 - 1	CCV87	27743	4.8858	-	7471B	PPB	27 Nov 2023 16:34:45	U	SMPL
CCB87 - 1	CCB87	53	-0.0141	-	7471B	PPB	27 Nov 2023 16:37:08	U	SMPL