



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128460

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	11/27/2023	13:30
2	O5546-01		1	14	100.00	>212.0	>212.0	11/27/2023	14:00
3	O5546-01DUP		1	14	100.00	>212.0	>212.0	11/27/2023	14:30
4	O5546-02		1	12	100.00	>212.0	>212.0	11/27/2023	15:00
5	O5546-03		1	14	100.00	>212.0	>212.0	11/27/2023	15:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB128460

WorkList Name : FP 11-27

WorkList ID : 175851

Department : Wet-Chemistry

Date : 11-27-2023 09:18:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5546-01	1145	Water	Flash Point	Cool 4 deg C	PSEG03	L31	11/27/2023	1010B
O5546-02	1146	Water	Flash Point	Cool 4 deg C	PSEG03	L31	11/27/2023	1010B
O5546-03	1147	Water	Flash Point	Cool 4 deg C	PSEG03	L31	11/27/2023	1010B

Date/Time 11/27/2023 12:00
 Raw Sample Received by: RPL (LWC)
 Raw Sample Relinquished by: 12/20/23

Date/Time 11/27/2023 16:10
 Raw Sample Received by: 12/20/23
 Raw Sample Relinquished by: RPL (LWC)