

LB128465


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37660



A= 0.0000e+000

B= 2.6631e-004 slope

C= -7.4480e-002 *intercept*

Rho= 0.9999198

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.075	-0.075	-3	0.000	-3					
0.05	0.050					419	0				
0.20	0.200	0.210	0.010	1068	0.0 %	1068					
2.50	2.500	2.555	0.055	9873	0.0 %	9873					
5.00	5.000	5.047	0.047	19232	0.0 %	19232					
7.50	7.500	7.508	0.008	28473	0.0 %	28473					
10.0	10.000	9.955	-0.045	37660	0.0 %	37660					

1152-00

LB128465

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-3	-		0 SFAM01.1	PPB	11/28/2023 7:24	Std
	0.2 S01	1068	-		0.2 SFAM01.1	PPB	11/28/2023 7:26	Std
	2.5 S02	9873	-		2.5 SFAM01.1	PPB	11/28/2023 7:29	Std
	5 S03	19232	-		5 SFAM01.1	PPB	11/28/2023 7:31	Std
	7.5 S04	28473	-		7.5 SFAM01.1	PPB	11/28/2023 7:33	Std
	10 S05	37660	-		10 SFAM01.1	PPB	11/28/2023 7:37	Std
ICV090	ICV090	15499	4.0531 -		SFAM01.1	PPB	11/28/2023 7:40	SMPL
ICB090	ICB090	231	-0.013 -		SFAM01.1	PPB	11/28/2023 7:45	SMPL
CCV088	CCV088	18482	4.8475 -		SFAM01.1	PPB	11/28/2023 7:47	SMPL
CCB088	CCB088	-75	-0.0945 -		SFAM01.1	PPB	11/28/2023 7:49	SMPL
PB157411BL	PBW411	-66	-0.0921 -		SFAM01.1	PPB	11/28/2023 7:52	SMPL
O5262-02	MDCKP7	-82	-0.0963 -		SFAM01.1	PPB	11/28/2023 7:54	SMPL
O5262-03	MDCKP8	-58	-0.0899 -		SFAM01.1	PPB	11/28/2023 7:56	SMPL
O5262-04	MDCKS2	-40	-0.0851 -		SFAM01.1	PPB	11/28/2023 7:59	SMPL
O5262-05	MDCKS2D	-23	-0.0806 -		SFAM01.1	PPB	11/28/2023 8:01	SMPL
O5262-06	MDCKS2S	4721	1.1828 -		SFAM01.1	PPB	11/28/2023 8:03	SMPL
O5262-07	MDCKP6	30201	7.9685 -		SFAM01.1	PPB	11/28/2023 8:05	SMPL
O5262-08	MDCKS0	-57	-0.0897 -		SFAM01.1	PPB	11/28/2023 8:11	SMPL
O5262-09	MDCKS1	-62	-0.091 -		SFAM01.1	PPB	11/28/2023 8:13	SMPL
O5262-10	MDCKS3	-28	-0.0819 -		SFAM01.1	PPB	11/28/2023 8:15	SMPL
O5262-11	MDCKS4	18	-0.0697 -		SFAM01.1	PPB	11/28/2023 8:17	SMPL
O5262-12	MDCKS5	78	-0.0537 -		SFAM01.1	PPB	11/28/2023 8:20	SMPL
O5262-13	MDCKS6	1753	0.3924 -		SFAM01.1	PPB	11/28/2023 8:22	SMPL
O5262-14	MDCKS7	-63	-0.0913 -		SFAM01.1	PPB	11/28/2023 8:24	SMPL
O5262-15	MDCKS8	1206	0.2467 -		SFAM01.1	PPB	11/28/2023 8:26	SMPL
O5262-16DLX10	MDCKS9	121	-0.0423 -		SFAM01.1	PPB	11/28/2023 8:29	SMPL
PB157412BL	PBW412	267	-0.0034 -		SFAM01.1	PPB	11/28/2023 8:31	SMPL
O5343-01	MCOAH1	493	0.0568 -		SFAM01.1	PPB	11/28/2023 8:33	SMPL
O5343-02	MCOAH1D	184	-0.0255 -		SFAM01.1	PPB	11/28/2023 8:35	SMPL
O5343-03	MCOAH1S	4471	1.1162 -		SFAM01.1	PPB	11/28/2023 8:40	SMPL
CCV089	CCV089	18880	4.9535 -		SFAM01.1	PPB	11/28/2023 8:45	SMPL
CCB089	CCB089	-4	-0.0755 -		SFAM01.1	PPB	11/28/2023 8:50	SMPL
O5344-01	MCOAH2	375	0.0254 -		SFAM01.1	PPB	11/28/2023 8:52	SMPL
O5344-02	MCOAH2D	-269	-0.1461 -		SFAM01.1	PPB	11/28/2023 8:54	SMPL
O5344-03	MCOAH2S	4100	1.0174 -		SFAM01.1	PPB	11/28/2023 9:05	SMPL
CCV090	CCV090	21544	5.663 -		SFAM01.1	PPB	11/28/2023 9:07	SMPL
CCB090	CCB090	129	-0.0401 -		SFAM01.1	PPB	11/28/2023 9:09	SMPL