

13128467


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

43749



A= 0.0000e+000

B= 2.2940e-004 *slope*

C= -5.2523e-002 *intercept*

Rho= 0.9999470

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>1/1</i>
0.00	0.000	-0.042	-0.042	48	0.000	48					
0.05	0.050					419	0				
0.20	0.200	0.185	-0.015	1035	0.0 %	1035					
2.50	2.500	2.551	0.051	11350	0.0 %	11350					
5.00	5.000	5.053	0.053	22257	0.0 %	22257					
7.50	7.500	7.469	-0.031	32786	0.0 %	32786					
10.0	10.000	9.984	-0.016	43749	0.0 %	43749					

118100

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	48	-		0 SFAM01.1	PPB	11/28/2023 10:35	Std
	0.2 S01	1035	-		0.2 SFAM01.1	PPB	11/28/2023 10:38	Std
	2.5 S02	11350	-		2.5 SFAM01.1	PPB	11/28/2023 10:40	Std
	5 S03	22257	-		5 SFAM01.1	PPB	11/28/2023 10:42	Std
	7.5 S04	32786	-		7.5 SFAM01.1	PPB	11/28/2023 10:44	Std
	10 S05	43749	-		10 SFAM01.1	PPB	11/28/2023 10:47	Std
ICV091	ICV091	17063	3.8617 -		SFAM01.1	PPB	11/28/2023 10:50	SMPL
ICB091	ICB091	62	-0.0383 -		SFAM01.1	PPB	11/28/2023 10:55	SMPL
CCV091	CCV091	22365	5.078 -		SFAM01.1	PPB	11/28/2023 10:57	SMPL
CCB091	CCB091	-177	-0.0931 -		SFAM01.1	PPB	11/28/2023 10:59	SMPL
PB157403BL	PBS403	-40	-0.0617 -		SFAM01.1	PPB	11/28/2023 11:02	SMPL
O5263-03	MDCKR5	748	0.1191 -		SFAM01.1	PPB	11/28/2023 11:04	SMPL
O5263-04	MDCKQ2	1896	0.3824 -		SFAM01.1	PPB	11/28/2023 11:06	SMPL
O5263-05	MDCKQ5	18695	4.2361 -		SFAM01.1	PPB	11/28/2023 11:08	SMPL
O5263-06	MDCKQ6	1186	0.2195 -		SFAM01.1	PPB	11/28/2023 11:11	SMPL
O5263-07	MDCKQ7	370	0.0324 -		SFAM01.1	PPB	11/28/2023 11:13	SMPL
O5263-08	MDCKQ8	251	0.0051 -		SFAM01.1	PPB	11/28/2023 11:15	SMPL
O5263-09	MDCKQ9	470	0.0553 -		SFAM01.1	PPB	11/28/2023 11:17	SMPL
O5263-10	MDCKR0	360	0.0301 -		SFAM01.1	PPB	11/28/2023 11:20	SMPL
O5263-11	MDCKR1	455	0.0519 -		SFAM01.1	PPB	11/28/2023 11:22	SMPL
O5263-12	MDCKR6	19910	4.5149 -		SFAM01.1	PPB	11/28/2023 11:24	SMPL
O5263-13	MDCKR7	15806	3.5734 -		SFAM01.1	PPB	11/28/2023 11:27	SMPL
O5263-14	MDCKR8	33860	7.715 -		SFAM01.1	PPB	11/28/2023 11:29	SMPL
O5263-15	MDCKT0	25052	5.6944 -		SFAM01.1	PPB	11/28/2023 11:31	SMPL
O5263-16	MDCKT5	478	0.0571 -		SFAM01.1	PPB	11/28/2023 11:33	SMPL
O5263-17	MDCKT5D	436	0.0475 -		SFAM01.1	PPB	11/28/2023 11:38	SMPL
O5263-18	MDCKT5S	9516	2.1305 -		SFAM01.1	PPB	11/28/2023 11:41	SMPL
CCV092	CCV092	19861	4.5036 -		SFAM01.1	PPB	11/28/2023 11:43	SMPL
CCB092	CCB092	-205	-0.0996 -		SFAM01.1	PPB	11/28/2023 11:45	SMPL