



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128468

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	11/28/2023	08:00
2	CAL2	1	Water	NA	NA	20.5	7.00	11/28/2023	08:01
3	CAL3	1	Water	NA	NA	20.5	10.02	11/28/2023	08:05
4	ICV	1	Water	NA	NA	20.6	7.01	11/28/2023	08:10
5	CCV1	1	Water	NA	NA	20.6	2.01	11/28/2023	08:14
6	O5548-02	1	Solid	20.02	20	22.9	5.67	11/28/2023	08:22
7	O5548-02DUP	1	Solid	20.03	20	22.8	5.68	11/28/2023	08:23
8	CCV2	1	Water	NA	NA	20.6	12.02	11/28/2023	08:25

128468

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROIVITY O5548 WorkList ID : 175855 Department : Wet-Chemistry Date : 11-28-2023 07:39:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5548-02	PAINT-PPE-PLASTICS	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	11/27/2023	9045D

Date/Time 11/28/23 07:50
Raw Sample Received by: SP WC
Raw Sample Relinquished by: SP WC

Date/Time 11/28/23 11:30
Raw Sample Received by: SP WC
Raw Sample Relinquished by: SP WC