



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128471

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	11/28/2023	08:40
2	O5547-01		1	12	100.00	>212.0	>212.0	11/28/2023	09:10
3	O5547-01DUP		1	12	100.00	>212.0	>212.0	11/28/2023	09:40
4	O5549-01		1	14	100.00	>212.0	>212.0	11/28/2023	10:10
5	O5550-01		1	14	100.00	>212.0	>212.0	11/28/2023	10:40
6	O5550-02		1	12	100.00	>212.0	>212.0	11/28/2023	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16128471

WorkList Name : FP-11-28 WorkList ID : 175861 Department : Wet-Chemistry Date : 11-28-2023 08:21:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5547-01	BCC-8002	Water	Flash Point	Cool 4 deg C	PSEG03	K31	11/27/2023	1010B
O5549-01	2861145	Water	Flash Point	Cool 4 deg C	PSEG03	L31	11/27/2023	1010B
O5550-01	A591	Water	Flash Point	Cool 4 deg C	PSEG03	L32	11/27/2023	1010B
O5550-02	A591	Water	Flash Point	Cool 4 deg C	PSEG03	L32	11/27/2023	1010B

Date/Time 11/28/2023 08:30
 Raw Sample Received by: RPA
 Raw Sample Relinquished by: Jd Coe

Date/Time 11/28/2023 11:45
 Raw Sample Received by: Jd Coe
 Raw Sample Relinquished by: RPA