

LB128478


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36472



A= 0.0000e+000

B= 2.7423e-004 *slope*C= -7.2473e-002 *y-intercept*

Rho= 0.9999155

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.049	-0.049	86	0.000	86					1
0.05	0.050	0.049	-0.001	442	0.0 %	442	0				-2
0.20	0.200	0.208	0.008	1022	0.0 %	1022					4
2.50	2.500	2.536	0.036	9513	0.0 %	9513					-1
5.00	5.000	4.991	-0.009	18464	0.0 %	18464					0
7.50	7.500	7.586	0.086	27928	0.0 %	27928					-1
10.0	10.000	9.929	-0.071	36472	0.0 %	36472					1

LB128473

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 SS0	86	-		0 SFAM01.1	PPB	11/28/2023 12:26	Std
	0.05 S01	442	-	0.05	SFAM01.1	PPB	11/28/2023 12:28	Std
	0.2 S02	1022	-	0.2	SFAM01.1	PPB	11/28/2023 12:30	Std
	2.5 S03	9513	-	2.5	SFAM01.1	PPB	11/28/2023 12:32	Std
	5 S04	18464	-	5	SFAM01.1	PPB	11/28/2023 12:35	Std
	7.5 S05	27928	-	7.5	SFAM01.1	PPB	11/28/2023 12:39	Std
	10 S06	36472	-	10	SFAM01.1	PPB	11/28/2023 12:52	Std
ICV092	ICV092	14490	3.9011 -		SFAM01.1	PPB	11/28/2023 12:56	SMPL
ICB092	ICB092	327	0.0172 -		SFAM01.1	PPB	11/28/2023 13:01	SMPL
CCV093	CCV093	17472	4.7188 -		SFAM01.1	PPB	11/28/2023 13:09	SMPL
CCB093	CCB093	400	0.0372 -		SFAM01.1	PPB	11/28/2023 13:13	SMPL
PB157445BL	PBW445	353	0.0243 -		SFAM01.1	PPB	11/28/2023 13:18	SMPL
O5327-01	EXEQ8	197	-0.0185 -		SFAM01.1	PPB	11/28/2023 13:20	SMPL
O5327-02	EXEQ9	-33	-0.0815 -		SFAM01.1	PPB	11/28/2023 13:23	SMPL
O5327-03	EXER1	24	-0.0659 -		SFAM01.1	PPB	11/28/2023 13:25	SMPL
O5327-04	EXER1D	53	-0.0579 -		SFAM01.1	PPB	11/28/2023 13:27	SMPL
O5327-05	EXER1S	3859	0.9858 -		SFAM01.1	PPB	11/28/2023 13:32	SMPL
CCV094	CCV094	19418	5.2525 -		SFAM01.1	PPB	11/28/2023 13:34	SMPL
CCB094	CCB094	102	-0.0445 -		SFAM01.1	PPB	11/28/2023 13:36	SMPL