

LB128479


SFAM01.1

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

46083



A= 0.0000e+000

B= 2.1785e-004 slope

C= -3.8937e-002 y-intercept

Rho= 0.9999467

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.00	0.000	-0.015	-0.015	111	0.000	111					
0.05	0.050					442	0				
0.20	0.200	0.235	0.035	1259	0.0 %	1259					
2.50	2.500	2.509	0.009	11695	0.0 %	11695					
5.00	5.000	4.927	-0.073	22796	0.0 %	22796					
7.50	7.500	7.543	0.043	34804	0.0 %	34804					
10.0	10.000	10.000	0.000	46083	0.0 %	46083					

96.11
11801
0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	111	-		0 SFAM01.1	PPB	11/28/2023 14:38	Std
	0.2 S01	1259	-		0.2 SFAM01.1	PPB	11/28/2023 14:40	Std
	2.5 S02	11695	-		2.5 SFAM01.1	PPB	11/28/2023 14:42	Std
	5 S03	22796	-		5 SFAM01.1	PPB	11/28/2023 14:45	Std
	7.5 S04	34804	-		7.5 SFAM01.1	PPB	11/28/2023 14:47	Std
	10 S05	46083	-		10 SFAM01.1	PPB	11/28/2023 14:52	Std
ICV093	ICV093	19270	4.159 -		SFAM01.1	PPB	11/28/2023 14:55	SMPL
ICB093	ICB093	-254	-0.0943 -		SFAM01.1	PPB	11/28/2023 14:57	SMPL
CCV095	CCV095	22823	4.9331 -		SFAM01.1	PPB	11/28/2023 14:59	SMPL
CCB095	CCB095	-40	-0.0477 -		SFAM01.1	PPB	11/28/2023 15:02	SMPL
PB157411BL	PBW411	10	-0.0368 -		SFAM01.1	PPB	11/28/2023 15:04	SMPL
O5262-02	MDCKP7	-52	-0.0503 -		SFAM01.1	PPB	11/28/2023 15:06	SMPL
O5262-03	MDCKP8	-38	-0.0472 -		SFAM01.1	PPB	11/28/2023 15:08	SMPL
O5262-04	MDCKS2	-3	-0.0396 -		SFAM01.1	PPB	11/28/2023 15:11	SMPL
O5262-05	MDCKS2D	-3	-0.0396 -		SFAM01.1	PPB	11/28/2023 15:13	SMPL
O5262-06	MDCKS2S	4916	1.032 -		SFAM01.1	PPB	11/28/2023 15:15	SMPL
O5262-07	MDCKP6	30862	6.6844 -		SFAM01.1	PPB	11/28/2023 15:17	SMPL
O5262-08	MDCKS0	-555	-0.1598 -		SFAM01.1	PPB	11/28/2023 15:20	SMPL
O5262-09	MDCKS1	-112	-0.0633 -		SFAM01.1	PPB	11/28/2023 15:22	SMPL
O5262-10	MDCKS3	-57	-0.0514 -		SFAM01.1	PPB	11/28/2023 15:24	SMPL
O5262-11	MDCKS4	13	-0.0361 -		SFAM01.1	PPB	11/28/2023 15:27	SMPL
O5262-12	MDCKS5	79	-0.0217 -		SFAM01.1	PPB	11/28/2023 15:29	SMPL
O5262-13	MDCKS6	1944	0.3846 -		SFAM01.1	PPB	11/28/2023 15:31	SMPL
O5262-14	MDCKS7	-69	-0.054 -		SFAM01.1	PPB	11/28/2023 15:33	SMPL
O5262-15	MDCKS8	1294	0.243 -		SFAM01.1	PPB	11/28/2023 15:36	SMPL
O5262-16DLX10	MDCKS9	297	0.0258 -		SFAM01.1	PPB	11/28/2023 15:38	SMPL
PB157412BL	PBW412	127	-0.0113 -		SFAM01.1	PPB	11/28/2023 15:40	SMPL
O5343-01	MCOAH1	545	0.0798 -		SFAM01.1	PPB	11/28/2023 15:42	SMPL
O5343-02	MCOAH1D	220	0.009 -		SFAM01.1	PPB	11/28/2023 15:45	SMPL
O5343-03	MCOAH1S	5220	1.0982 -		SFAM01.1	PPB	11/28/2023 15:47	SMPL
CCV096	CCV096	22591	4.8825 -		SFAM01.1	PPB	11/28/2023 15:52	SMPL
CCB096	CCB096	-7	-0.0405 -		SFAM01.1	PPB	11/28/2023 15:54	SMPL
O5344-01	MCOAH2	482	0.0661 -		SFAM01.1	PPB	11/28/2023 16:01	SMPL
O5344-02	MCOAH2D	413	0.051 -		SFAM01.1	PPB	11/28/2023 16:04	SMPL
O5344-03	MCOAH2S	5432	1.1444 -		SFAM01.1	PPB	11/28/2023 16:06	SMPL
CCV097	CCV097	21948	4.7424 -		SFAM01.1	PPB	11/28/2023 16:10	SMPL
CCB097	CCB097	-27	-0.0448 -		SFAM01.1	PPB	11/28/2023 16:13	SMPL