



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128484

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	11/28/2023	15:40
2	CAL2	1	Water	NA	NA	20.6	7.00	11/28/2023	15:42
3	CAL3	1	Water	NA	NA	20.7	10.02	11/28/2023	15:45
4	ICV	1	Water	NA	NA	20.5	7.01	11/28/2023	15:50
5	CCV1	1	Water	NA	NA	20.7	2.01	11/28/2023	15:55
6	O5556-01	1	Solid	20.02	20	22.6	9.20	11/28/2023	16:00
7	O5556-01DUP	1	Solid	20.03	20	22.7	9.21	11/28/2023	16:01
8	CCV2	1	Water	NA	NA	20.7	12.02	11/28/2023	16:05

VB 128484

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O5556 WorkList ID : 175877 Department : Wet-Chemistry Date : 11-28-2023 15:35:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5556-01	SP-1-COMP	Solid	pH	Cool 4 deg C	EART12	L11	11/28/2023	9045D

Date/Time 11/28/23 15430
Raw Sample Received by: JO/ACC/ 5055m
Raw Sample Relinquished by: JO/ACC/ 5055m

Date/Time 11/28/23 17110
Raw Sample Received by: JO/ACC/ 5055m
Raw Sample Relinquished by: JO/ACC/ 5055m